

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV100323\
 Data File : VV032210.D
 Acq On : 03 Oct 2023 11:09
 Operator : SY/MD
 Sample : VV1003WBL01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK205

Quant Time: Oct 04 04:05:28 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092523WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Oct 04 04:04:33 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	203675	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.789	117	197325	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	97563	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	75109	59.589	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	119.180%
7) Chloroethane-d5	1.532	69	61651	56.708	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	113.420%
11) 1,1-Dichloroethene-d2	2.060	65	31109	56.522	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	113.040%
21) 2-Butanone-d5	3.786	46	94647	83.511	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	83.510%
24) Chloroform-d	4.253	84	127092	48.184	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.360%
26) 1,2-Dichloroethane-d4	4.944	65	78231	49.295	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.580%
32) Benzene-d6	4.963	84	259858	50.095	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.180%
36) 1,2-Dichloropropane-d6	5.992	67	86310	49.860	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	99.720%
41) Toluene-d8	7.246	98	233165	49.271	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.540%
43) trans-1,3-Dichloroprop...	7.558	79	36375	44.172	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	88.340%
47) 2-Hexanone-d5	8.027	63	57557	76.997	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	77.000%
56) 1,1,2,2-Tetrachloroeth...	10.156	84	106426	43.227	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	86.460%
66) 1,2-Dichlorobenzene-d4	11.564	152	88646	50.103	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	100.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV100323\
Data File : VV032210.D
Acq On : 03 Oct 2023 11:09
Operator : SY/MD
Sample : VV1003WBL01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK205

Quant Time: Oct 04 04:05:28 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092523WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Oct 04 04:04:33 2023
Response via : Initial Calibration

