

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV100324\
 Data File : VV037725.D
 Acq On : 03 Oct 2024 10:47
 Operator : SY/MD
 Sample : P4237-16ME
 Misc : 4.86G/5mL/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 A4EN9ME

Quant Time: Oct 04 02:28:05 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Oct 04 02:26:43 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.526	114	848596	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.776	117	820798	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.178	152	403014	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.281	65	228506	30.675	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	61.340%
7) Chloroethane-d5	1.507	69	82591	14.284	ug/L	-0.03
Spiked Amount	50.000	Range	70 - 130	Recovery	=	28.560%#
11) 1,1-Dichloroethene-d2	2.043	65	104727	33.650	ug/L	-0.02
Spiked Amount	50.000	Range	60 - 125	Recovery	=	67.300%
21) 2-Butanone-d5	3.799	46	186996	70.477	ug/L	0.01
Spiked Amount	100.000	Range	40 - 130	Recovery	=	70.480%
24) Chloroform-d	4.240	84	457013	34.483	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	68.960%#
26) 1,2-Dichloroethane-d4	4.934	65	282846	36.461	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	72.920%
32) Benzene-d6	4.947	84	907975	35.351	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	70.700%
36) 1,2-Dichloropropane-d6	5.982	67	297543	36.380	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	72.760%
41) Toluene-d8	7.236	98	800226	34.359	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	68.720%#
43) trans-1,3-Dichloroprop...	7.548	79	126808	33.377	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	66.760%
47) 2-Hexanone-d5	8.027	63	123415	71.462	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	71.460%
56) 1,1,2,2-Tetrachloroeth...	10.146	84	326887	34.398	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	68.800%
66) 1,2-Dichlorobenzene-d4	11.554	152	329902	38.847	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	77.700%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV100324\
Data File : VV037725.D
Acq On : 03 Oct 2024 10:47
Operator : SY/MD
Sample : P4237-16ME
Misc : 4.86G/5mL/100uL/5.0mL/MSVOA_V/MEOH
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
A4EN9ME

Quant Time: Oct 04 02:28:05 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092624WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Oct 04 02:26:43 2024
Response via : Initial Calibration

