

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV100325\
 Data File : VV039250.D
 Acq On : 03 Oct 2025 15:27
 Operator : SY/MD
 Sample : Q3201-08
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 TRIP-BLANK

Quant Time: Oct 04 01:12:02 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V092225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 24 08:08:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	4.603	168	323029	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	5.535	114	535676	50.000	ug/l	0.00
63) Chlorobenzene-d5	8.777	117	545144	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.175	152	264732	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.944	65	259998	55.612	ug/l	0.00
Spiked Amount	50.000	Range	81 - 118	Recovery	=	111.220%
35) Dibromofluoromethane	4.503	113	254125	59.011	ug/l	0.00
Spiked Amount	50.000	Range	80 - 119	Recovery	=	118.020%
50) Toluene-d8	7.240	98	559880	44.266	ug/l	0.00
Spiked Amount	50.000	Range	89 - 112	Recovery	=	88.540%#
62) 4-Bromofluorobenzene	10.002	95	232498	44.653	ug/l	0.00
Spiked Amount	50.000	Range	85 - 114	Recovery	=	89.300%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV100325\
Data File : VV039250.D
Acq On : 03 Oct 2025 15:27
Operator : SY/MD
Sample : Q3201-08
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
TRIP-BLANK

Quant Time: Oct 04 01:12:02 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V092225W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 24 08:08:08 2025
Response via : Initial Calibration

