

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV100325\
 Data File : VV039256.D
 Acq On : 03 Oct 2025 17:41
 Operator : SY/MD
 Sample : Q3201-01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 MW6

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 10/06/2025
 Supervised By :Semsettin Yesilyurt 10/06/2025

Quant Time: Oct 04 01:14:44 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V092225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 24 08:08:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	4.600	168	338723	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	5.535	114	596501	50.000	ug/l	0.00
63) Chlorobenzene-d5	8.776	117	602719	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.175	152	285289	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.944	65	295630	60.304	ug/l	0.00
Spiked Amount 50.000	Range 81 - 118		Recovery = 120.600%#			
35) Dibromofluoromethane	4.503	113	276319	57.622	ug/l	0.00
Spiked Amount 50.000	Range 80 - 119		Recovery = 115.240%			
50) Toluene-d8	7.240	98	631507	44.838	ug/l	0.00
Spiked Amount 50.000	Range 89 - 112		Recovery = 89.680%			
62) 4-Bromofluorobenzene	10.002	95	253846	43.781	ug/l	0.00
Spiked Amount 50.000	Range 85 - 114		Recovery = 87.560%			
Target Compounds						Qvalue
11) Tert butyl alcohol	2.564	59	7717590	7448.531	ug/l	99
16) Acetone	2.124	43	24177	3.247	ug/l	99
19) Methyl tert-butyl Ether	2.690	73	39182020m	2407.667	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV100325\
Data File : VV039256.D
Acq On : 03 Oct 2025 17:41
Operator : SY/MD
Sample : Q3201-01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
MW6

Quant Time: Oct 04 01:14:44 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V092225W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 24 08:08:08 2025
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 10/06/2025
Supervised By :Semsettin Yesilyurt 10/06/2025

