

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV100325\
 Data File : VV039258.D
 Acq On : 03 Oct 2025 18:26
 Operator : SY/MD
 Sample : Q3201-03
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 MW15

Quant Time: Oct 04 01:15:41 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V092225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 24 08:08:08 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	4.600	168	410821	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	5.532	114	719975	50.000	ug/l	0.00
63) Chlorobenzene-d5	8.776	117	749276	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.175	152	364005	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.944	65	352717	59.322	ug/l	0.00
Spiked Amount	50.000	Range	81 - 118	Recovery	= 118.640%#	
35) Dibromofluoromethane	4.500	113	343112	59.280	ug/l	0.00
Spiked Amount	50.000	Range	80 - 119	Recovery	= 118.560%	
50) Toluene-d8	7.239	98	780643	45.922	ug/l	0.00
Spiked Amount	50.000	Range	89 - 112	Recovery	= 91.840%	
62) 4-Bromofluorobenzene	10.001	95	320017	45.728	ug/l	0.00
Spiked Amount	50.000	Range	85 - 114	Recovery	= 91.460%	
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.577	59	11991	9.542	ug/l #	91
16) Acetone	2.127	43	18201	0.349	ug/l	94
17) Carbon Disulfide	2.256	76	5502	0.290	ug/l	98
19) Methyl tert-butyl Ether	2.722	73	13805	0.699	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV100325\
Data File : VV039258.D
Acq On : 03 Oct 2025 18:26
Operator : SY/MD
Sample : Q3201-03
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
MW15

Quant Time: Oct 04 01:15:41 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V092225W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 24 08:08:08 2025
Response via : Initial Calibration

