

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_V\DATA\VV100418\
 Data File : VV008079.D
 Acq On : 04 Oct 2018 23:19
 Operator : SY/MD
 Sample : J5246-13
 Misc : 25.0 mL/MSVOA_V/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0J81

Quant Time: Oct 05 06:51:09 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR092718WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Oct 05 05:56:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	78541	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	75693	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	36314	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	17621	4.18	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	83.60%
7) Chloroethane-d5	1.59	69	17221	4.73	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	94.60%
11) 1,1-Dichloroethene-d2	2.14	63	29784	3.13	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	62.60%
20) 2-Butanone-d5	3.96	46	72692	57.40	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	114.80%
24) Chloroform-d	4.41	84	48965	5.11	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.20%
26) 1,2-Dichloroethane-d4	5.09	65	26575	5.04	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.80%
32) Benzene-d6	5.10	84	90436	5.01	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.20%
36) 1,2-Dichloropropane-d6	6.13	67	29546	5.16	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	103.20%
41) Toluene-d8	7.36	98	80161	4.36	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.20%
43) trans-1,3-Dichloropropene-	7.67	79	9912	4.49	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	89.80%
46) 2-Hexanone-d5	8.14	63	45785	47.02	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	94.04%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	22166	5.01	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	100.20%
64) 1,2-Dichlorobenzene-d4	11.68	152	35708	5.08	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	101.60%

Target Compounds

					Qvalue
18) trans-1,2-Dichloroethene	2.79	96	855	0.149 ug/L	90
19) 1,1-Dichloroethane	3.23	63	1151	0.118 ug/L #	84
22) cis-1,2-Dichloroethene	3.97	96	14417	2.488 ug/L	95
34) Trichloroethene	5.96	95	25348	4.084 ug/L	96
47) Tetrachloroethene	8.03	164	769	0.134 ug/L #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_V\DATA\VV100418\
Data File : VV008079.D
Acq On : 04 Oct 2018 23:19
Operator : SY/MD
Sample : J5246-13
Misc : 25.0 mL/MSVOA_V/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0J81

Quant Time: Oct 05 06:51:09 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR092718WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Oct 05 05:56:33 2018
Response via : Initial Calibration

