

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_V\DATA\VV100418\
 Data File : VV008081.D
 Acq On : 05 Oct 2018 00:14
 Operator : SY/MD
 Sample : J5246-15
 Misc : 25.0 mL/MSVOA_V/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0J83

Quant Time: Oct 05 06:57:03 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR092718WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Oct 05 05:56:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	80322	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	77517	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	39365	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	17419	4.04	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	80.80%
7) Chloroethane-d5	1.59	69	16668	4.48	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	89.60%
11) 1,1-Dichloroethene-d2	2.13	63	28019	2.88	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	57.60%#
20) 2-Butanone-d5	3.98	46	75551	58.33	ug/L	0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	116.66%
24) Chloroform-d	4.41	84	49237	5.02	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.40%
26) 1,2-Dichloroethane-d4	5.09	65	25525	4.73	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.60%
32) Benzene-d6	5.10	84	90510	4.89	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.80%
36) 1,2-Dichloropropane-d6	6.12	67	30413	5.18	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	103.60%
41) Toluene-d8	7.36	98	80523	4.27	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	85.40%
43) trans-1,3-Dichloropropene-	7.67	79	9074	4.02	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	80.40%
46) 2-Hexanone-d5	8.14	63	50125	50.27	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	100.54%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	23858	5.27	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	105.40%
64) 1,2-Dichlorobenzene-d4	11.68	152	35338	4.64	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	92.80%

Target Compounds

					Qvalue
14) Carbon disulfide	2.32	76	9355	0.650 ug/L	99
22) cis-1,2-Dichloroethene	3.97	96	1507	0.254 ug/L	90
34) Trichloroethene	5.96	95	25475	4.008 ug/L	94
42) Toluene	7.43	91	2585	0.107 ug/L	93
47) Tetrachloroethene	8.02	164	24816	4.214 ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_V\DATA\VV100418\
Data File : VV008081.D
Acq On : 05 Oct 2018 00:14
Operator : SY/MD
Sample : J5246-15
Misc : 25.0 mL/MSVOA_V/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0J83

Quant Time: Oct 05 06:57:03 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR092718WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Oct 05 05:56:33 2018
Response via : Initial Calibration

