

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_V\DATA\VV100418\
 Data File : VV008083.D
 Acq On : 05 Oct 2018 01:09
 Operator : SY/MD
 Sample : J5246-17
 Misc : 25.0 mL/MSVOA_V/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0J85

Quant Time: Oct 05 07:06:03 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR092718WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Oct 05 05:56:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	79236	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	75613	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	36932	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	16767	3.95	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	79.00%
7) Chloroethane-d5	1.59	69	16034	4.37	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	87.40%
11) 1,1-Dichloroethene-d2	2.14	63	30326	3.16	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	63.20%
20) 2-Butanone-d5	3.97	46	70294	55.02	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	110.04%
24) Chloroform-d	4.41	84	48703	5.03	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.60%
26) 1,2-Dichloroethane-d4	5.09	65	24767	4.65	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.00%
32) Benzene-d6	5.10	84	87937	4.87	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.40%
36) 1,2-Dichloropropane-d6	6.13	67	29812	5.21	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	104.20%
41) Toluene-d8	7.36	98	78153	4.25	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	85.00%
43) trans-1,3-Dichloropropene-	7.67	79	9351	4.24	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	84.80%
46) 2-Hexanone-d5	8.14	63	46662	47.98	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	95.96%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	22970	5.20	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	104.00%
64) 1,2-Dichlorobenzene-d4	11.68	152	34464	4.82	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	96.40%

Target Compounds

					Qvalue
12) 1,1-Dichloroethene	2.14	96	4245	0.811 ug/L	# 1
22) cis-1,2-Dichloroethene	3.97	96	11734	2.007 ug/L	91
25) Chloroform	4.43	83	9292	0.872 ug/L	96
31) Carbon tetrachloride	4.88	117	63184	7.883 ug/L	99
34) Trichloroethene	5.96	95	48374	7.803 ug/L	94
47) Tetrachloroethene	8.02	164	39941	6.952 ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_V\DATA\VV100418\
Data File : VV008083.D
Acq On : 05 Oct 2018 01:09
Operator : SY/MD
Sample : J5246-17
Misc : 25.0 mL/MSVOA_V/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0J85

Quant Time: Oct 05 07:06:03 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR092718WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Oct 05 05:56:33 2018
Response via : Initial Calibration

