

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_V\DATA\VV100418\
 Data File : VV008055.D
 Acq On : 04 Oct 2018 11:17
 Operator : SY/MD
 Sample : VSTDCCC005
 Misc : 25.0 mL/MSVOA_V/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_V
 LabSampleId :
 VSTD00570

Quant Time: Oct 05 05:18:15 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR092718WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Oct 04 07:54:45 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 200%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Difluorobenzene	5.000	5.000	0.0	102	0.00
2 T	Dichlorodifluoromethane	5.000	5.064	-1.3	104	0.00
3 T	Chloromethane	5.000	4.689	6.2	96	0.00
4 S	Vinyl Chloride-d3	5.000	4.311	13.8	89	0.00
5 T	Vinyl chloride	5.000	4.702	6.0	96	0.00
6 T	Bromomethane	5.000	4.682	6.4	96	0.00
7 S	Chloroethane-d5	5.000	4.420	11.6	92	0.00
8 T	Chloroethane	5.000	4.754	4.9	98	0.00
9 T	Trichlorofluoromethane	5.000	4.901	2.0	100	0.00
10 T	1,1,2-Trichloro-1,2,2-trifl	5.000	5.076	-1.5	104	0.00
11 S	1,1-Dichloroethene-d2	5.000	4.407	11.9	93	0.00
12 T	1,1-Dichloroethene	5.000	4.789	4.2	98	0.00
13 T	Acetone	50.000	44.934	10.1	99	0.00
14 T	Carbon disulfide	5.000	4.323	13.5	93	0.00
15 T	Methyl Acetate	5.000	4.424	11.5	104	0.00
16 T	Methylene chloride	5.000	4.647	7.1	102	0.00
17 T	Methyl tert-butyl Ether	5.000	4.777	4.5	101	0.00
18 T	trans-1,2-Dichloroethene	5.000	4.717	5.7	99	0.00
19 T	1,1-Dichloroethane	5.000	5.125	-2.5	103	0.00
20 S	2-Butanone-d5	50.000	49.739	0.5	108	0.00
21 T	2-Butanone	50.000	46.904	6.2	103	0.00
22 T	cis-1,2-Dichloroethene	5.000	5.264	-5.3	108	0.00
23 T	Bromochloromethane	5.000	5.457	-9.1	112	0.00
24 S	Chloroform-d	5.000	4.884	2.3	101	0.00
25 T	Chloroform	5.000	5.225	-4.5	107	0.00
26 S	1,2-Dichloroethane-d4	5.000	4.578	8.4	97	0.00
27 T	1,2-Dichloroethane	5.000	4.790	4.2	98	0.00
28 I	Chlorobenzene-d5	5.000	5.000	0.0	101	0.00
29 T	1,1,1-Trichloroethane	5.000	5.354	-7.1	110	0.00
30 T	Cyclohexane	5.000	4.795	4.1	99	0.00
31 T	Carbon tetrachloride	5.000	5.429	-8.6	111	0.00
32 S	Benzene-d6	5.000	4.822	3.6	100	0.00
33 T	Benzene	5.000	5.160	-3.2	106	0.00
34 T	Trichloroethene	5.000	5.026	-0.5	105	0.00
35 T	Methylcyclohexane	5.000	4.871	2.6	101	0.00
36 S	1,2-Dichloropropane-d6	5.000	4.794	4.1	101	0.00
37 T	1,2-Dichloropropane	5.000	5.277	-5.5	111	0.00
38 T	Bromodichloromethane	5.000	5.458	-9.2	113	0.00
39 T	cis-1,3-Dichloropropene	5.000	5.283	-5.7	111	0.00
40 T	4-Methyl-2-pentanone	50.000	48.722	2.6	104	0.00
41 S	Toluene-d8	5.000	4.736	5.3	97	0.00
42 T	Toluene	5.000	5.251	-5.0	107	0.00
43 S	trans-1,3-Dichloropropene-d	5.000	4.892	2.2	106	0.00
44 T	trans-1,3-Dichloropropene	5.000	5.287	-5.7	109	0.00
45 T	1,1,2-Trichloroethane	5.000	5.059	-1.2	108	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_V\DATA\VV100418\
 Data File : VV008055.D
 Acq On : 04 Oct 2018 11:17
 Operator : SY/MD
 Sample : VSTDCCC005
 Misc : 25.0 mL/MSVOA_V/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_V
 LabSampleId :
 VSTD00570

Quant Time: Oct 05 05:18:15 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR092718WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Oct 04 07:54:45 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 200%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
46 S	2-Hexanone-d5	50.000	47.157	5.7	101	0.00
47 T	Tetrachloroethene	5.000	5.145	-2.9	107	0.00
48 T	2-Hexanone	50.000	49.526	0.9	103	0.00
49 T	Dibromochloromethane	5.000	5.614	-12.3	117	0.00
50 T	1,2-Dibromoethane	5.000	5.165	-3.3	105	0.00
51 T	Chlorobenzene	5.000	5.180	-3.6	108	0.00
52 T	Ethylbenzene	5.000	5.116	-2.3	105	0.00
53 T	m,p-xylene	5.000	5.104	-2.1	105	0.00
54 T	o-xylene	5.000	5.109	-2.2	105	0.00
55 T	Styrene	5.000	5.246	-4.9	106	0.00
56 T	Isopropylbenzene	5.000	5.218	-4.4	106	0.00
57 S	1,1,2,2-Tetrachloroethane-d	5.000	4.762	4.8	101	0.00
58 T	1,1,2,2-Tetrachloroethane	5.000	5.029	-0.6	106	0.00
59	1,2,3-Trichloropropane	5.000	5.066	-1.3	107	0.00
60 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	99	0.00
61 T	Bromoform	5.000	5.786	-15.7	124	0.00
62 T	1,3-Dichlorobenzene	5.000	5.152	-3.0	107	0.00
63 T	1,4-Dichlorobenzene	5.000	5.067	-1.3	106	0.00
64 S	1,2-Dichlorobenzene-d4	5.000	4.442	11.2	97	0.00
65 T	1,2-Dichlorobenzene	5.000	5.291	-5.8	110	0.00
66 T	1,2-Dibromo-3-chloropropane	5.000	5.339	-6.8	109	0.00
67	1,3,5-Trichlorobenzene	5.000	5.245	-4.9	109	0.00
68 T	1,2,4-trichlorobenzene	5.000	4.715	5.7	101	0.00
69	Naphthalene	5.000	4.110	17.8	86	0.00
70 T	1,2,3-Trichlorobenzene	5.000	4.929	1.4	102	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0