

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_V\DATA\VV100418\
 Data File : VV008086.D
 Acq On : 05 Oct 2018 02:31
 Operator : SY/MD
 Sample : VSTDCCC005EC
 Misc : 25.0 mL/MSVOA_V/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_V
 LabSampleId :
 VSTD00572

Quant Time: Oct 05 06:00:30 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR092718WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Oct 05 05:56:33 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 200%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Difluorobenzene	5.000	5.000	0.0	89	0.00
2 T	Dichlorodifluoromethane	5.000	5.017	-0.3	91	0.00
3 T	Chloromethane	5.000	4.534	9.3	81	0.00
4 S	Vinyl Chloride-d3	5.000	4.189	16.2	76	0.00
5 T	Vinyl chloride	5.000	5.037	-0.7	90	0.00
6 T	Bromomethane	5.000	2.047	59.1#	37	0.00
7 S	Chloroethane-d5	5.000	4.554	8.9	83	0.00
8 T	Chloroethane	5.000	5.234	-4.7	95	0.00
9 T	Trichlorofluoromethane	5.000	5.123	-2.5	92	0.00
10 T	1,1,2-Trichloro-1,2,2-trifl	5.000	5.277	-5.5	95	0.00
11 S	1,1-Dichloroethene-d2	5.000	4.418	11.6	81	0.00
12 T	1,1-Dichloroethene	5.000	5.109	-2.2	92	0.00
13 T	Acetone	50.000	47.836	4.3	92	0.00
14 T	Carbon disulfide	5.000	4.722	5.6	89	0.00
15 T	Methyl Acetate	5.000	5.003	-0.1	103	0.00
16 T	Methylene chloride	5.000	4.984	0.3	96	0.00
17 T	Methyl tert-butyl Ether	5.000	5.222	-4.4	97	0.00
18 T	trans-1,2-Dichloroethene	5.000	5.188	-3.8	95	0.00
19 T	1,1-Dichloroethane	5.000	5.576	-11.5	98	0.00
20 S	2-Butanone-d5	50.000	56.031	-12.1	107	0.00
21 T	2-Butanone	50.000	52.961	-5.9	102	0.00
22 T	cis-1,2-Dichloroethene	5.000	5.615	-12.3	101	0.00
23 T	Bromochloromethane	5.000	5.831	-16.6	105	0.00
24 S	Chloroform-d	5.000	5.312	-6.2	96	0.00
25 T	Chloroform	5.000	5.745	-14.9	103	0.00
26 S	1,2-Dichloroethane-d4	5.000	4.733	5.3	88	0.00
27 T	1,2-Dichloroethane	5.000	5.358	-7.2	96	0.00
28 I	Chlorobenzene-d5	5.000	5.000	0.0	87	0.00
29 T	1,1,1-Trichloroethane	5.000	5.780	-15.6	102	0.00
30 T	Cyclohexane	5.000	5.068	-1.4	90	0.00
31 T	Carbon tetrachloride	5.000	5.876	-17.5	103	0.00
32 S	Benzene-d6	5.000	5.105	-2.1	91	0.00
33 T	Benzene	5.000	5.740	-14.8	101	0.00
34 T	Trichloroethene	5.000	5.826	-16.5	105	0.00
35 T	Methylcyclohexane	5.000	4.981	0.4	88	0.00
36 S	1,2-Dichloropropane-d6	5.000	5.348	-7.0	96	0.00
37 T	1,2-Dichloropropane	5.000	5.988	-19.8	108	0.00
38 T	Bromodichloromethane	5.000	6.145	-22.9	109	0.00
39 T	cis-1,3-Dichloropropene	5.000	5.490	-9.8	98	0.00
40 T	4-Methyl-2-pentanone	50.000	55.945	-11.9	102	0.00
41 S	Toluene-d8	5.000	4.942	1.2	86	0.00
42 T	Toluene	5.000	5.771	-15.4	101	0.00
43 S	trans-1,3-Dichloropropene-d	5.000	4.783	4.3	89	0.00
44 T	trans-1,3-Dichloropropene	5.000	5.510	-10.2	98	0.00
45 T	1,1,2-Trichloroethane	5.000	5.796	-15.9	106	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_V\DATA\VV100418\
 Data File : VV008086.D
 Acq On : 05 Oct 2018 02:31
 Operator : SY/MD
 Sample : VSTDCCC005EC
 Misc : 25.0 mL/MSVOA_V/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_V
 LabSampleId :
 VSTD00572

Quant Time: Oct 05 06:00:30 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR092718WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Oct 05 05:56:33 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 200%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
46 S	2-Hexanone-d5	50.000	54.245	-8.5	100	0.00
47 T	Tetrachloroethene	5.000	5.331	-6.6	95	0.00
48 T	2-Hexanone	50.000	57.820	-15.6	103	0.00
49 T	Dibromochloromethane	5.000	6.328	-26.6	113	0.00
50 T	1,2-Dibromoethane	5.000	5.845	-16.9	102	0.00
51 T	Chlorobenzene	5.000	5.575	-11.5	100	0.00
52 T	Ethylbenzene	5.000	5.461	-9.2	96	0.00
53 T	m,p-xylene	5.000	5.584	-11.7	99	0.00
54 T	o-xylene	5.000	5.549	-11.0	98	0.00
55 T	Styrene	5.000	5.716	-14.3	99	0.00
56 T	Isopropylbenzene	5.000	5.602	-12.0	97	0.00
57 S	1,1,2,2-Tetrachloroethane-d	5.000	5.489	-9.8	100	0.00
58 T	1,1,2,2-Tetrachloroethane	5.000	5.932	-18.6	107	0.00
59	1,2,3-Trichloropropane	5.000	5.713	-14.3	103	0.00
60 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	88	0.00
61 T	Bromoform	5.000	6.186	-23.7	118	0.00
62 T	1,3-Dichlorobenzene	5.000	5.389	-7.8	99	0.00
63 T	1,4-Dichlorobenzene	5.000	5.285	-5.7	99	0.00
64 S	1,2-Dichlorobenzene-d4	5.000	4.638	7.2	90	0.00
65 T	1,2-Dichlorobenzene	5.000	5.494	-9.9	101	0.00
66 T	1,2-Dibromo-3-chloropropane	5.000	5.662	-13.2	103	0.00
67	1,3,5-Trichlorobenzene	5.000	5.316	-6.3	98	0.00
68 T	1,2,4-trichlorobenzene	5.000	4.967	0.7	94	0.00
69	Naphthalene	5.000	4.576	8.5	85	0.00
70 T	1,2,3-Trichlorobenzene	5.000	5.249	-5.0	96	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0