

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV100423\
 Data File : VV032226.D
 Acq On : 04 Oct 2023 13:53
 Operator : SY/MD
 Sample : 04708-06
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DCJT2

Quant Time: Oct 05 01:41:18 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR092923WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Oct 05 01:37:52 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.538	114	98002	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.789	117	95891	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.191	152	46485	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	37292	4.679	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	93.600%
7) Chloroethane-d5	1.532	69	32768	5.065	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	101.200%
11) 1,1-Dichloroethene-d2	2.060	65	15761	4.689	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	93.800%
20) 2-Butanone-d5	3.805	46	95655	47.442	ug/L	-0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	94.880%
24) Chloroform-d	4.252	84	74866	5.325	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	106.400%
26) 1,2-Dichloroethane-d4	4.950	65	32820	4.928	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.600%
32) Benzene-d6	4.963	84	137541	4.875	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.400%
36) 1,2-Dichloropropane-d6	5.995	67	46083	4.929	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	98.600%
41) Toluene-d8	7.249	98	117153	4.541	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.800%
43) trans-1,3-Dichloroprop...	7.558	79	15290	4.616	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	92.400%
46) 2-Hexanone-d5	8.030	63	73830	47.368	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	94.740%
56) 1,1,2,2-Tetrachloroeth...	10.156	84	31991	4.893	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	97.800%
66) 1,2-Dichlorobenzene-d4	11.564	152	43566	5.051	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	101.000%
Target Compounds						
					Qvalue	
13) Acetone	2.130	43	4408	2.837	ug/L	99
25) Chloroform	4.278	83	171283	11.272	ug/L	99
34) Trichloroethene	5.834	95	833559	90.443	ug/L	98
38) Bromodichloromethane	6.442	83	10061	0.938	ug/L	99
47) Tetrachloroethene	7.911	164	2267	0.324	ug/L	92
49) Dibromochloromethane	8.185	129	1395	0.210	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV100423\
Data File : VV032226.D
Acq On : 04 Oct 2023 13:53
Operator : SY/MD
Sample : 04708-06
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DCJT2

Quant Time: Oct 05 01:41:18 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR092923WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Oct 05 01:37:52 2023
Response via : Initial Calibration

