

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV100423\
 Data File : VV032231.D
 Acq On : 04 Oct 2023 15:53
 Operator : SY/MD
 Sample : 04741-01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AZ1

Quant Time: Oct 05 01:42:28 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR092923WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Oct 05 01:37:52 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.539	114	101484	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.789	117	97574	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.191	152	46237	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	35677	4.323	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	86.400%	
7) Chloroethane-d5	1.532	69	30775	4.593	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	91.800%	
11) 1,1-Dichloroethene-d2	2.060	65	14949	4.295	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	85.800%	
20) 2-Butanone-d5	3.806	46	96209	46.080	ug/L	-0.01
Spiked Amount	50.000	Range 40 - 130	Recovery	=	92.160%	
24) Chloroform-d	4.256	84	65332	4.487	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	89.800%	
26) 1,2-Dichloroethane-d4	4.950	65	33184	4.812	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	96.200%	
32) Benzene-d6	4.963	84	130827	4.557	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	91.200%	
36) 1,2-Dichloropropane-d6	5.992	67	44862	4.716	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	94.400%	
41) Toluene-d8	7.249	98	115248	4.390	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	87.800%	
43) trans-1,3-Dichloroprop...	7.561	79	14814	4.395	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	87.800%	
46) 2-Hexanone-d5	8.031	63	70595	44.511	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	89.020%	
56) 1,1,2,2-Tetrachloroeth...	10.159	84	32022	4.813	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	96.200%	
66) 1,2-Dichlorobenzene-d4	11.564	152	42237	4.923	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	98.400%	
Target Compounds						Qvalue
25) Chloroform	4.281	83	14859	0.944	ug/L	96
38) Bromodichloromethane	6.439	83	20147	1.846	ug/L	97
47) Tetrachloroethene	7.912	164	11218	1.577	ug/L	97
49) Dibromochloromethane	8.182	129	13826	2.049	ug/L	95
59) Bromoform	9.674	173	2501	0.744	ug/L #	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV100423\
Data File : VV032231.D
Acq On : 04 Oct 2023 15:53
Operator : SY/MD
Sample : 04741-01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AZ1

Quant Time: Oct 05 01:42:28 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR092923WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Oct 05 01:37:52 2023
Response via : Initial Calibration

