

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV100518\
 Data File : VV008097.D
 Acq On : 05 Oct 2018 14:48
 Operator : SY/MD
 Sample : J5246-06RE
 Misc : 25.0 mL/MSVOA V/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0J74RE

Quant Time: Oct 06 01:15:25 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR100518WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Oct 06 01:03:16 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	125848	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	121650	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	58526	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	31892	4.13	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	82.60%
7) Chloroethane-d5	1.59	69	27020	4.15	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	83.00%
11) 1,1-Dichloroethene-d2	2.14	63	48308	3.19	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	63.80%
20) 2-Butanone-d5	3.96	46	80560	43.35	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	86.70%
24) Chloroform-d	4.41	84	65460	4.10	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	82.00%
26) 1,2-Dichloroethane-d4	5.09	65	33142	4.16	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	83.20%
32) Benzene-d6	5.10	84	130084	4.19	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	83.80%
36) 1,2-Dichloropropane-d6	6.12	67	40287	4.09	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	81.80%
41) Toluene-d8	7.36	98	119662	3.96	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	79.20%
43) trans-1,3-Dichloropropene-	7.67	79	13933	4.06	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	81.20%
46) 2-Hexanone-d5	8.14	63	56895	41.59	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	83.18%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	27209	3.93	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	78.60%
64) 1,2-Dichlorobenzene-d4	11.68	152	48457	4.36	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	87.20%

Target Compounds

					Ovalue
13) Acetone	2.22	43	2431	1.931 ug/L	97
22) cis-1,2-Dichloroethene	3.97	96	5050	0.602 ug/L	95
34) Trichloroethene	5.96	95	6313	0.705 ug/L	98
47) Tetrachloroethene	8.02	164	1617	0.202 ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV100518\
Data File : VV008097.D
Acq On : 05 Oct 2018 14:48
Operator : SY/MD
Sample : J5246-06RE
Misc : 25.0 mL/MSVOA V/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0J74RE

Quant Time: Oct 06 01:15:25 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR100518WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Oct 06 01:03:16 2018
Response via : Initial Calibration

