

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV100518\
 Data File : VV008098.D
 Acq On : 05 Oct 2018 15:15
 Operator : SY/MD
 Sample : J5246-11RE
 Misc : 25.0 mL/MSVOA V/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0J79RE

Quant Time: Oct 08 07:15:35 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR100518WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Oct 06 01:03:16 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	110806	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	105852	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	50018	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	26725	3.93	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	78.60%
7) Chloroethane-d5	1.59	69	23130	4.03	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	80.60%
11) 1,1-Dichloroethene-d2	2.14	63	41541	3.12	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	62.40%
20) 2-Butanone-d5	3.96	46	69367	42.39	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	84.78%
24) Chloroform-d	4.41	84	56190	4.00	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	80.00%
26) 1,2-Dichloroethane-d4	5.09	65	28738	4.10	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	82.00%
32) Benzene-d6	5.10	84	107815	3.99	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	79.80%
36) 1,2-Dichloropropane-d6	6.12	67	33918	3.95	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	79.00%
41) Toluene-d8	7.36	98	100067	3.81	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	76.20%
43) trans-1,3-Dichloropropene-	7.67	79	11588	3.88	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	77.60%
46) 2-Hexanone-d5	8.14	63	49304	41.42	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	82.84%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	23286	3.87	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	77.40%
64) 1,2-Dichlorobenzene-d4	11.68	152	39772	4.19	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	83.80%

Target Compounds

					Ovalue
13) Acetone	2.21	43	2832	2.554	ug/L 91
22) cis-1,2-Dichloroethene	3.96	96	2135	0.289	ug/L 92
31) Carbon tetrachloride	4.88	117	2922	0.277	ug/L 97
34) Trichloroethene	5.96	95	10759	1.380	ug/L 98
47) Tetrachloroethene	8.02	164	6528	0.938	ug/L 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV100518\
Data File : VV008098.D
Acq On : 05 Oct 2018 15:15
Operator : SY/MD
Sample : J5246-11RE
Misc : 25.0 mL/MSVOA V/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0J79RE

Quant Time: Oct 08 07:15:35 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR100518WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Oct 06 01:03:16 2018
Response via : Initial Calibration

