

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV100518\
 Data File : VV008131.D
 Acq On : 06 Oct 2018 06:27
 Operator : SY/MD
 Sample : J5277-08
 Misc : 25.0 mL/MSVOA V/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BEYC5

Quant Time: Oct 08 05:37:14 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR100518WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Mon Oct 08 05:22:34 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	97777	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	92969	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	42552	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	24546	4.09	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	81.80%
7) Chloroethane-d5	1.59	69	23076	4.56	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	91.20%
11) 1,1-Dichloroethene-d2	2.14	63	37112	3.16	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	63.20%
20) 2-Butanone-d5	3.96	46	72650	50.31	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	100.62%
24) Chloroform-d	4.41	84	56010	4.52	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.40%
26) 1,2-Dichloroethane-d4	5.09	65	29508	4.77	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.40%
32) Benzene-d6	5.10	84	106589	4.49	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.80%
36) 1,2-Dichloropropane-d6	6.12	67	34786	4.62	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	92.40%
41) Toluene-d8	7.36	98	93502	4.05	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	81.00%
43) trans-1,3-Dichloropropene-	7.67	79	10594	4.04	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	80.80%
46) 2-Hexanone-d5	8.14	63	45859	43.87	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	87.74%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	24347	4.61	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	92.20%
64) 1,2-Dichlorobenzene-d4	11.68	152	38376	4.75	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	95.00%

Target Compounds					Ovalue
13) Acetone	2.21	43	2205	2.254 ug/L	96
34) Trichloroethene	5.96	95	82867	12.105 ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV100518\
Data File : VV008131.D
Acq On : 06 Oct 2018 06:27
Operator : SY/MD
Sample : J5277-08
Misc : 25.0 mL/MSVOA V/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BEYC5

Quant Time: Oct 08 05:37:14 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR100518WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Mon Oct 08 05:22:34 2018
Response via : Initial Calibration

