

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV100518\
 Data File : VV008127.D
 Acq On : 06 Oct 2018 04:38
 Operator : SY/MD
 Sample : J5277-04
 Misc : 25.0 mL/MSVOA V/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BEYC1

Quant Time: Oct 06 06:55:11 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR100518WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Oct 06 04:00:29 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	96967	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	95115	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	43036	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	25269	4.25	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	85.00%
7) Chloroethane-d5	1.59	69	22366	4.45	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	89.00%
11) 1,1-Dichloroethene-d2	2.14	63	37960	3.26	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	65.20%
20) 2-Butanone-d5	3.96	46	70430	49.18	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	98.36%
24) Chloroform-d	4.41	84	55767	4.54	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.80%
26) 1,2-Dichloroethane-d4	5.09	65	28655	4.67	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.40%
32) Benzene-d6	5.10	84	106599	4.39	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.80%
36) 1,2-Dichloropropane-d6	6.12	67	34909	4.53	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	90.60%
41) Toluene-d8	7.36	98	91175	3.86	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	77.20%
43) trans-1,3-Dichloropropene-	7.67	79	10050	3.74	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	74.80%
46) 2-Hexanone-d5	8.14	63	48588	45.43	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	90.86%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	24039	4.45	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	89.00%
64) 1,2-Dichlorobenzene-d4	11.68	152	39107	4.79	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	95.80%


Target Compounds					Ovalue
13) Acetone	2.21	43	1887	1.945 ug/L	93


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV100518\
Data File : VV008127.D
Acq On : 06 Oct 2018 04:38
Operator : SY/MD
Sample : J5277-04
Misc : 25.0 mL/MSVOA V/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BEYC1

Quant Time: Oct 06 06:55:11 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR100518WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Oct 06 04:00:29 2018
Response via : Initial Calibration





#13
 Acetone
 Concen: 1.945 ug/L
 RT: 2.21 min Scan# 348
 Delta R.T. -0.00 min
 Lab File: VV008127.D
 Acq: 06 Oct 2018 04:38

Instrument :
 MSVOA_V
 ClientSampleId :
 BEYC1

Tot Ion: 43 Resp: 1887
 Ion Ratio Lower Upper
 43 100
 58 27.4 0.0 62.2

