

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV100520\
 Data File : VV018680.D
 Acq On : 05 Oct 2020 17:32
 Operator : SY/MD
 Sample : L4254-18
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BG3W4

Quant Time: Oct 06 03:19:52 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM092920WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Oct 06 02:11:36 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	242312	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	237936	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	108828	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	71985	50.65	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	101.30%
7) Chloroethane-d5	1.58	69	59783	50.88	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	101.76%
11) 1,1-Dichloroethene-d2	2.12	63	117997	40.75	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	81.50%
21) 2-Butanone-d5	3.91	46	81903	77.26	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	77.26%
24) Chloroform-d	4.37	84	127970	41.64	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	83.28%
26) 1,2-Dichloroethane-d4	5.05	65	89606	45.93	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.86%
32) Benzene-d6	5.07	84	246567	40.91	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	81.82%
36) 1,2-Dichloropropane-d6	6.09	67	76077	41.78	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	83.56%
41) Toluene-d8	7.34	98	226136	39.88	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	79.76%#
43) trans-1,3-Dichloropropene-	7.64	79	39237	39.89	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	79.78%
47) 2-Hexanone-d5	8.11	63	40369	58.36	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	58.36%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	92694	38.39	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	76.78%
64) 1,2-Dichlorobenzene-d4	11.65	152	91412	43.68	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	87.36%


Target Compounds					Ovalue
13) Acetone	2.21	43	4967	5.424 ug/L	67


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV100520\
Data File : VV018680.D
Acq On : 05 Oct 2020 17:32
Operator : SY/MD
Sample : L4254-18
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BG3W4

Quant Time: Oct 06 03:19:52 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM092920WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Oct 06 02:11:36 2020
Response via : Initial Calibration





#13
 Acetone
 Concen: 5.424 ug/L
 RT: 2.21 min Scan# 347
 Delta R.T. 0.00 min
 Lab File: VV018680.D
 Acq: 05 Oct 2020 17:32

Instrument :
 MSVOA_V
 ClientSampleId :
 BG3W4

Tot Ion: 43 Resp: 4967
 Ion Ratio Lower Upper
 43 100
 58 13.7 0.0 64.0

