

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW100521\
 Data File : VW022590.D
 Acq On : 06 Oct 2021 00:12
 Operator : SY/MD
 Sample : M4051-05
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AT3

Quant Time: Oct 06 08:06:59 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR100421WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Oct 06 07:50:25 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	98687	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	99473	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.252	152	44625	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	37399	11.828	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	= 236.600%#		
7) Chloroethane-d5	1.564	69	32588	7.646	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	= 153.000%#		
11) 1,1-Dichloroethene-d2	2.105	63	56235	6.130	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	= 122.600%		
20) 2-Butanone-d5	3.880	46	93567	50.061	ug/L	-0.03
Spiked Amount	50.000	Range 40 - 130	Recovery	= 100.120%		
24) Chloroform-d	4.349	84	71373	5.862	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 117.200%		
26) 1,2-Dichloroethane-d4	5.034	65	36118	6.297	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 126.000%		
32) Benzene-d6	5.050	84	142098	6.339	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 126.800%#		
36) 1,2-Dichloropropane-d6	6.072	67	43320	5.478	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	= 109.600%		
41) Toluene-d8	7.317	98	122075	6.496	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 130.000%		
43) trans-1,3-Dichloroprop...	7.625	79	12412	5.930	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	= 118.600%		
46) 2-Hexanone-d5	8.088	63	57356	44.260	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	= 88.520%		
56) 1,1,2,2-Tetrachloroeth...	10.217	84	28144	4.684	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	= 93.600%		
66) 1,2-Dichlorobenzene-d4	11.625	152	47077	6.820	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	= 136.400%#		
Target Compounds						
						Qvalue
13) Acetone	2.169	43	3549	3.531	ug/L	# 38
14) Carbon disulfide	2.291	76	24109	1.363	ug/L	98
34) Trichloroethene	5.921	95	457	0.059	ug/L	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV100521\
Data File : VV022590.D
Acq On : 06 Oct 2021 00:12
Operator : SY/MD
Sample : M4051-05
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AT3

Quant Time: Oct 06 08:06:59 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR100421WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Oct 06 07:50:25 2021
Response via : Initial Calibration

