

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW100521\
 Data File : VW022572.D
 Acq On : 05 Oct 2021 15:02
 Operator : SY/MD
 Sample : M3967-11
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Oct 06 05:28:21 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR100421WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Oct 06 05:22:36 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	105671	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	100815	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.252	152	46813	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	15232	4.499	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	90.000%
7) Chloroethane-d5	1.564	69	22546	4.940	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	98.800%
11) 1,1-Dichloroethene-d2	2.105	63	30094	3.063	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	61.200%
20) 2-Butanone-d5	3.915	46	91297	45.618	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	91.240%
24) Chloroform-d	4.349	84	57696	4.425	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.600%
26) 1,2-Dichloroethane-d4	5.034	65	30260	4.927	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.600%
32) Benzene-d6	5.047	84	104116	4.583	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.600%
36) 1,2-Dichloropropane-d6	6.072	67	35999	4.491	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	89.800%
41) Toluene-d8	7.317	98	80574	4.231	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	84.600%
43) trans-1,3-Dichloroprop...	7.628	79	9727	4.586	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	91.800%
46) 2-Hexanone-d5	8.095	63	64785	49.327	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	98.660%
56) 1,1,2,2-Tetrachloroeth...	10.220	84	28416	4.666	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	93.400%
66) 1,2-Dichlorobenzene-d4	11.628	152	37247	5.144	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	102.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV100521\
Data File : VV022572.D
Acq On : 05 Oct 2021 15:02
Operator : SY/MD
Sample : M3967-11
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Oct 06 05:28:21 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR100421WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Oct 06 05:22:36 2021
Response via : Initial Calibration

