

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV100523\
 Data File : VV032262.D
 Acq On : 05 Oct 2023 17:47
 Operator : SY/MD
 Sample : 04748-07
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EX882

Quant Time: Oct 06 01:31:04 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR092923WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Oct 06 01:20:04 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	5.538	114	79409	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.789	117	83697	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.191	152	38920	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	29740	4.605	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	92.000%	
7) Chloroethane-d5	1.532	69	28067	5.354	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	107.000%	
11) 1,1-Dichloroethene-d2	2.060	65	12480	4.582	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	91.600%	
20) 2-Butanone-d5	3.802	46	86657	53.043	ug/L	-0.02
Spiked Amount	50.000	Range 40 - 130	Recovery	=	106.080%	
24) Chloroform-d	4.256	84	60057	5.271	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	105.400%	
26) 1,2-Dichloroethane-d4	4.950	65	29728	5.509	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	110.200%	
32) Benzene-d6	4.966	84	114209	4.637	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	92.800%	
36) 1,2-Dichloropropane-d6	5.992	67	42721	5.236	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	104.800%	
41) Toluene-d8	7.249	98	103354	4.589	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	91.800%	
43) trans-1,3-Dichloroprop...	7.561	79	13302	4.600	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	92.000%	
46) 2-Hexanone-d5	8.030	63	69422	51.029	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	102.060%	
56) 1,1,2,2-Tetrachloroeth...	10.159	84	30175	5.287	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	105.800%	
66) 1,2-Dichlorobenzene-d4	11.564	152	38576	5.341	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	106.800%	
Target Compounds						
22) cis-1,2-Dichloroethene	3.825	96	1055	0.147	ug/L	96
25) Chloroform	4.281	83	31376	2.548	ug/L	96
29) 1,1,1-Trichloroethane	4.516	97	4534	0.398	ug/L	94
31) Carbon tetrachloride	4.738	117	46493	4.829	ug/L	99
34) Trichloroethene	5.834	95	659122	81.936	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV100523\
Data File : VV032262.D
Acq On : 05 Oct 2023 17:47
Operator : SY/MD
Sample : 04748-07
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
EX882

Quant Time: Oct 06 01:31:04 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR092923WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Oct 06 01:20:04 2023
Response via : Initial Calibration

