

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW100523\
 Data File : VW032274.D
 Acq On : 05 Oct 2023 23:01
 Operator : SY/MD
 Sample : 04741-07
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AZ7

Quant Time: Oct 06 05:27:11 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR092923WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Oct 06 05:25:10 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.539	114	86034	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.789	117	92916	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	42748	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	27542	3.936	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	78.800%	
7) Chloroethane-d5	1.532	69	26224	4.617	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	92.400%	
11) 1,1-Dichloroethene-d2	2.060	65	11281	3.823	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	76.400%	
20) 2-Butanone-d5	3.809	46	83097	46.947	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	93.900%	
24) Chloroform-d	4.253	84	56084	4.544	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	90.800%	
26) 1,2-Dichloroethane-d4	4.947	65	27737	4.744	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	94.800%	
32) Benzene-d6	4.963	84	107633	3.937	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	78.800%	
36) 1,2-Dichloropropane-d6	5.992	67	39249	4.333	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	86.600%	
41) Toluene-d8	7.249	98	94719	3.789	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	75.800%	
43) trans-1,3-Dichloroprop...	7.561	79	12102	3.770	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	75.400%	
46) 2-Hexanone-d5	8.031	63	60421	40.006	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	80.020%	
56) 1,1,2,2-Tetrachloroeth...	10.156	84	27900	4.403	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	88.000%	
66) 1,2-Dichlorobenzene-d4	11.564	152	36092	4.550	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	91.000%	
Target Compounds						Qvalue
25) Chloroform	4.278	83	16681	1.250	ug/L	97
38) Bromodichloromethane	6.439	83	22231	2.139	ug/L	97
47) Tetrachloroethene	7.912	164	10508	1.552	ug/L	99
49) Dibromochloromethane	8.182	129	15374	2.393	ug/L	96
59) Bromoform	9.670	173	2794	0.899	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV100523\
Data File : VV032274.D
Acq On : 05 Oct 2023 23:01
Operator : SY/MD
Sample : 04741-07
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AZ7

Quant Time: Oct 06 05:27:11 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR092923WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Oct 06 05:25:10 2023
Response via : Initial Calibration

