

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV100620\
 Data File : VV018688.D
 Acq On : 06 Oct 2020 12:25
 Operator : SY/MD
 Sample : VV1006WBL05
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK307

Quant Time: Oct 07 06:11:06 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SFAMVLM092320W.M
 Quant Title : VOC Analysis
 QLast Update : Wed Oct 07 05:40:49 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	226762	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	251590	50.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.27	152	100024	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	71056	50.50	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	101.00%
7) Chloroethane-d5	1.58	69	60432	49.90	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	99.80%
11) 1,1-Dichloroethene-d2	2.12	63	110275	37.38	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	74.76%
21) 2-Butanone-d5	3.91	46	92141	91.35	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	91.35%
24) Chloroform-d	4.37	84	135084	45.78	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.56%
26) 1,2-Dichloroethane-d4	5.05	65	98106	52.98	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	105.96%
32) Benzene-d6	5.07	84	261459	41.71	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	83.42%
36) 1,2-Dichloropropane-d6	6.09	67	81848	40.07	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	80.14%
41) Toluene-d8	7.34	98	238026	41.05	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	82.10%
43) trans-1,3-Dichloropropene-	7.64	79	43076	42.02	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	84.04%
47) 2-Hexanone-d5	8.11	63	52561	69.90	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	69.90%
56) 1,1,2,2-Tetrachloroethane-	10.24	84	88566	33.76	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	67.52%
66) 1,2-Dichlorobenzene-d4	11.65	152	98097	52.68	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	105.36%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV100620\
Data File : VV018688.D
Acq On : 06 Oct 2020 12:25
Operator : SY/MD
Sample : VV1006WBL05
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK307

Quant Time: Oct 07 06:11:06 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SFAMVLM092320W.M
Quant Title : VOC Analysis
QLast Update : Wed Oct 07 05:40:49 2020
Response via : Initial Calibration

