

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\Data\VV100621\  
 Data File : VV022611.D  
 Acq On : 06 Oct 2021 10:57  
 Operator : SY/MD  
 Sample : VV1006WBL01  
 Misc : 25.0mL/MSVOA\_V/WATER  
 ALS Vial : 4 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 VBLK215

Quant Time: Oct 07 06:16:35 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVTR100421WMA.M  
 Quant Title : TRACE VOA SFAM1.0  
 QLast Update : Thu Oct 07 06:15:29 2021  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon  | Response | Conc     | Units | Dev(Min) |
|-------------------------------|--------|-------|----------|----------|-------|----------|
| -----                         |        |       |          |          |       |          |
| Internal Standards            |        |       |          |          |       |          |
| 1) 1,4-Difluorobenzene        | 5.619  | 114   | 100344   | 5.000    | ug/L  | 0.00     |
| 28) Chlorobenzene-d5          | 8.854  | 117   | 99865    | 5.000    | ug/L  | 0.00     |
| 58) 1,4-Dichlorobenzene-d4    | 11.252 | 152   | 45650    | 5.000    | ug/L  | 0.00     |
| System Monitoring Compounds   |        |       |          |          |       |          |
| 4) Vinyl Chloride-d3          | 1.304  | 65    | 17359    | 5.399    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 40 - 130 | Recovery | =     | 108.000% |
| 7) Chloroethane-d5            | 1.565  | 69    | 26114    | 6.026    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 65 - 130 | Recovery | =     | 120.600% |
| 11) 1,1-Dichloroethene-d2     | 2.105  | 63    | 34447    | 3.693    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 60 - 125 | Recovery | =     | 73.800%  |
| 20) 2-Butanone-d5             | 3.902  | 46    | 63995    | 33.674   | ug/L  | 0.00     |
| Spiked Amount                 | 50.000 | Range | 40 - 130 | Recovery | =     | 67.340%  |
| 24) Chloroform-d              | 4.349  | 84    | 61399    | 4.959    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 70 - 125 | Recovery | =     | 99.200%  |
| 26) 1,2-Dichloroethane-d4     | 5.034  | 65    | 30429    | 5.218    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 70 - 130 | Recovery | =     | 104.400% |
| 32) Benzene-d6                | 5.050  | 84    | 110437   | 4.908    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 70 - 125 | Recovery | =     | 98.200%  |
| 36) 1,2-Dichloropropane-d6    | 6.072  | 67    | 37545    | 4.729    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 60 - 140 | Recovery | =     | 94.600%  |
| 41) Toluene-d8                | 7.320  | 98    | 87682    | 4.648    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 70 - 130 | Recovery | =     | 93.000%  |
| 43) trans-1,3-Dichloroprop... | 7.625  | 79    | 9166     | 4.362    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 55 - 130 | Recovery | =     | 87.200%  |
| 46) 2-Hexanone-d5             | 8.092  | 63    | 57853    | 44.468   | ug/L  | 0.00     |
| Spiked Amount                 | 50.000 | Range | 45 - 130 | Recovery | =     | 88.940%  |
| 56) 1,1,2,2-Tetrachloroeth... | 10.220 | 84    | 27974    | 4.637    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 65 - 120 | Recovery | =     | 92.800%  |
| 66) 1,2-Dichlorobenzene-d4    | 11.625 | 152   | 40035    | 5.670    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 80 - 120 | Recovery | =     | 113.400% |

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\Data\VV100621\  
Data File : VV022611.D  
Acq On : 06 Oct 2021 10:57  
Operator : SY/MD  
Sample : VV1006WBL01  
Misc : 25.0mL/MSVOA\_V/WATER  
ALS Vial : 4 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
VBLK215

Quant Time: Oct 07 06:16:35 2021  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVTR100421WMA.M  
Quant Title : TRACE VOA SFAM1.0  
QLast Update : Thu Oct 07 06:15:29 2021  
Response via : Initial Calibration

