

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV100621\
 Data File : VV022639.D
 Acq On : 06 Oct 2021 23:02
 Operator : SY/MD
 Sample : M4050-07
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 GB7C3

Quant Time: Oct 07 07:47:22 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR100421WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Oct 07 07:45:03 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	101180	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	102270	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.252	152	46590	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	13658	4.213	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	84.200%
7) Chloroethane-d5	1.565	69	21695	4.965	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	99.200%
11) 1,1-Dichloroethene-d2	2.105	63	27725	2.948	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	59.000%#
20) 2-Butanone-d5	3.902	46	71725	37.429	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	74.860%
24) Chloroform-d	4.352	84	56919	4.560	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.200%
26) 1,2-Dichloroethane-d4	5.034	65	29641	5.041	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.800%
32) Benzene-d6	5.053	84	102632	4.453	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.000%
36) 1,2-Dichloropropane-d6	6.072	67	35366	4.350	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	87.000%
41) Toluene-d8	7.320	98	76996	3.985	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	79.800%
43) trans-1,3-Dichloroprop...	7.629	79	8772	4.077	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	81.600%
46) 2-Hexanone-d5	8.092	63	60905	45.713	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	91.420%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	27604	4.468	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	89.400%
66) 1,2-Dichlorobenzene-d4	11.625	152	37405	5.190	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	103.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV100621\
Data File : VV022639.D
Acq On : 06 Oct 2021 23:02
Operator : SY/MD
Sample : M4050-07
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
GB7C3

Quant Time: Oct 07 07:47:22 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR100421WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Oct 07 07:45:03 2021
Response via : Initial Calibration

