

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV100625\
 Data File : VV039266.D
 Acq On : 06 Oct 2025 12:13
 Operator : SY/MD
 Sample : VV1006MBL01
 Misc : 5.00g/10.0mL/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VV1006MBL01

Quant Time: Oct 07 03:31:10 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V092225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 24 08:08:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	4.600	168	491791	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	5.532	114	1015770	50.000	ug/l	0.00
63) Chlorobenzene-d5	8.776	117	966866	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.175	152	445450	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.940	65	421917	59.277	ug/l	0.00
Spiked Amount	50.000	Range	81 - 118	Recovery	=	118.560%#
35) Dibromofluoromethane	4.500	113	381093	46.668	ug/l	0.00
Spiked Amount	50.000	Range	80 - 119	Recovery	=	93.340%
50) Toluene-d8	7.236	98	1216997	50.743	ug/l	0.00
Spiked Amount	50.000	Range	89 - 112	Recovery	=	101.480%
62) 4-Bromofluorobenzene	10.001	95	397285	40.238	ug/l	0.00
Spiked Amount	50.000	Range	85 - 114	Recovery	=	80.480%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV100625\
Data File : VV039266.D
Acq On : 06 Oct 2025 12:13
Operator : SY/MD
Sample : VV1006MBL01
Misc : 5.00g/10.0mL/100uL/5.0mL/MSVOA_V/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VV1006MBL01

Quant Time: Oct 07 03:31:10 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V092225W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 24 08:08:08 2025
Response via : Initial Calibration

