

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV100625\
 Data File : VV039267.D
 Acq On : 06 Oct 2025 12:35
 Operator : SY/MD
 Sample : VV1006WBL01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VV1006WBL01

Quant Time: Oct 07 03:31:41 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V092225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 24 08:08:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	4.600	168	463806	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	5.535	114	783914	50.000	ug/l	0.00
63) Chlorobenzene-d5	8.776	117	804136	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.172	152	428242	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.944	65	340870	50.780	ug/l	0.00
Spiked Amount	50.000	Range	81 - 118	Recovery	=	101.560%
35) Dibromofluoromethane	4.503	113	339151	53.816	ug/l	0.00
Spiked Amount	50.000	Range	80 - 119	Recovery	=	107.640%
50) Toluene-d8	7.239	98	796316	43.023	ug/l	0.00
Spiked Amount	50.000	Range	89 - 112	Recovery	=	86.040%#
62) 4-Bromofluorobenzene	10.001	95	368553	48.368	ug/l	0.00
Spiked Amount	50.000	Range	85 - 114	Recovery	=	96.740%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV100625\
Data File : VV039267.D
Acq On : 06 Oct 2025 12:35
Operator : SY/MD
Sample : VV1006WBL01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VV1006WBL01

Quant Time: Oct 07 03:31:41 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V092225W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 24 08:08:08 2025
Response via : Initial Calibration

