

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV100625\
 Data File : VV039271.D
 Acq On : 06 Oct 2025 14:19
 Operator : SY/MD
 Sample : Q3201-01DL 100X
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 MW6DL

Quant Time: Oct 07 03:34:39 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V092225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 24 08:08:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	4.600	168	477065	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	5.532	114	810750	50.000	ug/l	0.00
63) Chlorobenzene-d5	8.776	117	821368	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.175	152	441845	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.940	65	354921	51.404	ug/l	0.00
Spiked Amount	50.000	Range	81 - 118	Recovery	=	102.800%
35) Dibromofluoromethane	4.503	113	354025	54.317	ug/l	0.00
Spiked Amount	50.000	Range	80 - 119	Recovery	=	108.640%
50) Toluene-d8	7.236	98	818283	42.746	ug/l	0.00
Spiked Amount	50.000	Range	89 - 112	Recovery	=	85.500%#
62) 4-Bromofluorobenzene	10.001	95	376058	47.720	ug/l	0.00
Spiked Amount	50.000	Range	85 - 114	Recovery	=	95.440%
Target Compounds						Qvalue
11) Tert butyl alcohol	2.571	59	82225	56.346	ug/l	97
19) Methyl tert-butyl Ether	2.712	73	944490	41.207	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV100625\
Data File : VV039271.D
Acq On : 06 Oct 2025 14:19
Operator : SY/MD
Sample : Q3201-01DL 100X
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
MW6DL

Quant Time: Oct 07 03:34:39 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V092225W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 24 08:08:08 2025
Response via : Initial Calibration

