

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV100625\
 Data File : VV039276.D
 Acq On : 06 Oct 2025 16:11
 Operator : SY/MD
 Sample : Q3289-02
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VNJ-240

Quant Time: Oct 07 03:37:08 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V092225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 24 08:08:08 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	4.603	168	152342	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	5.532	114	316837	50.000	ug/l	0.00
63) Chlorobenzene-d5	8.776	117	328845	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.175	152	141807	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.947	65	131671	59.719	ug/l	0.00
Spiked Amount	50.000	Range 81 - 118	Recovery	= 119.440%#		
35) Dibromofluoromethane	4.507	113	133337	52.348	ug/l	0.00
Spiked Amount	50.000	Range 80 - 119	Recovery	= 104.700%		
50) Toluene-d8	7.240	98	322905	43.164	ug/l	0.00
Spiked Amount	50.000	Range 89 - 112	Recovery	= 86.320%#		
62) 4-Bromofluorobenzene	10.005	95	130127	42.253	ug/l	0.00
Spiked Amount	50.000	Range 85 - 114	Recovery	= 84.500%#		
Target Compounds						
					Qvalue	
16) Acetone	2.124	43	72108	46.760	ug/l	96
18) Methyl Acetate	2.384	43	20589	6.302	ug/l	93
20) Methylene Chloride	2.455	84	26094	9.144	ug/l	93
43) Isopropyl Acetate	5.214	43	33727	5.294	ug/l #	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV100625\
Data File : VV039276.D
Acq On : 06 Oct 2025 16:11
Operator : SY/MD
Sample : Q3289-02
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VNJ-240

Quant Time: Oct 07 03:37:08 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V092225W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 24 08:08:08 2025
Response via : Initial Calibration

