

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV100625\
 Data File : VV039279.D
 Acq On : 06 Oct 2025 17:18
 Operator : SY/MD
 Sample : VIBLK
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK

Quant Time: Oct 07 03:38:41 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V092225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 24 08:08:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	4.638	168	565	50.000	ug/l	# 0.04
34) 1,4-Difluorobenzene	5.587	114	248	50.000	ug/l	0.06
63) Chlorobenzene-d5	8.815	117	1356	50.000	ug/l	0.04
72) 1,4-Dichlorobenzene-d4	11.201	152	1386	50.000	ug/l	0.03

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.982	65	634	77.532	ug/l	0.04
Spiked Amount	50.000	Range	81 - 118	Recovery	=	155.060%#
35) Dibromofluoromethane	4.516	113	256	128.403	ug/l	0.02
Spiked Amount	50.000	Range	80 - 119	Recovery	=	256.800%#
50) Toluene-d8	7.313	98	238	40.645	ug/l	0.08
Spiked Amount	50.000	Range	89 - 112	Recovery	=	81.280%#
62) 4-Bromofluorobenzene	10.034	95	368	152.660	ug/l	0.04
Spiked Amount	50.000	Range	85 - 114	Recovery	=	305.320%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV100625\
Data File : VV039279.D
Acq On : 06 Oct 2025 17:18
Operator : SY/MD
Sample : VIBLK
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK

Quant Time: Oct 07 03:38:41 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V092225W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 24 08:08:08 2025
Response via : Initial Calibration

