

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV100721\
 Data File : VV022672.D
 Acq On : 07 Oct 2021 19:05
 Operator : SY/MD
 Sample : M4051-21
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Oct 08 04:03:54 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR100721WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Oct 08 04:01:59 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	229669	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	223485	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.252	152	101287	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	76412	4.836	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	96.800%
7) Chloroethane-d5	1.564	69	69738	4.911	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	98.200%
11) 1,1-Dichloroethene-d2	2.101	63	115375	3.722	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	74.400%
20) 2-Butanone-d5	3.921	46	186134	44.925	ug/L	-0.04
Spiked Amount	50.000	Range	40 - 130	Recovery	=	89.840%
24) Chloroform-d	4.349	84	130210	3.999	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	80.000%
26) 1,2-Dichloroethane-d4	5.034	65	77209	4.885	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.800%
32) Benzene-d6	5.047	84	317510	4.772	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.400%
36) 1,2-Dichloropropane-d6	6.072	67	96681	4.907	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	98.200%
41) Toluene-d8	7.317	98	271201	4.715	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.400%
43) trans-1,3-Dichloroprop...	7.625	79	30667	4.692	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	93.800%
46) 2-Hexanone-d5	8.098	63	137688	52.274	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	104.540%
56) 1,1,2,2-Tetrachloroeth...	10.220	84	63548	4.761	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	95.200%
66) 1,2-Dichlorobenzene-d4	11.625	152	108079	5.755	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	115.200%
Target Compounds					Qvalue	
25) Chloroform	4.375	83	31104	0.987	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV100721\
Data File : VV022672.D
Acq On : 07 Oct 2021 19:05
Operator : SY/MD
Sample : M4051-21
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Oct 08 04:03:54 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR100721WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Oct 08 04:01:59 2021
Response via : Initial Calibration

