

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV100818\
 Data File : VV008146.D
 Acq On : 08 Oct 2018 12:00
 Operator : SY/MD
 Sample : J5277-19
 Misc : 25.0 mL/MSVOA V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BEYD6

Quant Time: Oct 09 08:42:06 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR100518WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Oct 09 03:37:28 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	105735	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	102207	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	47141	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	25651	3.95	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	79.00%
7) Chloroethane-d5	1.59	69	20731	3.79	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	75.80%
11) 1,1-Dichloroethene-d2	2.13	63	37800	2.97	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	59.40%#
20) 2-Butanone-d5	3.96	46	59618	38.18	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	76.36%
24) Chloroform-d	4.41	84	51406	3.84	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	76.80%
26) 1,2-Dichloroethane-d4	5.09	65	25922	3.87	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	77.40%
32) Benzene-d6	5.10	84	100174	3.84	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	76.80%
36) 1,2-Dichloropropane-d6	6.12	67	30888	3.73	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	74.60%
41) Toluene-d8	7.36	98	90900	3.58	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	71.60%
43) trans-1,3-Dichloropropene-	7.67	79	10967	3.80	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	76.00%
46) 2-Hexanone-d5	8.14	63	38312	33.34	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	66.68%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	20764	3.57	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	71.40%
64) 1,2-Dichlorobenzene-d4	11.68	152	35831	4.00	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	80.00%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV100818\
Data File : VV008146.D
Acq On : 08 Oct 2018 12:00
Operator : SY/MD
Sample : J5277-19
Misc : 25.0 mL/MSVOA V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BEYD6

Quant Time: Oct 09 08:42:06 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR100518WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Oct 09 03:37:28 2018
Response via : Initial Calibration

