

Data Path : Z:\voasrv\HPCHEM1\MSVOA V\Data\VV100818\
 Data File : VV008162.D
 Acq On : 08 Oct 2018 19:45
 Operator : SY/MD
 Sample : J5281-06
 Misc : 25.0 mL/MSVOA V/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BEYL3

Quant Time: Oct 15 19:11:13 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR100518WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Oct 09 04:07:48 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	83800	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	82594	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	37625	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	25563	4.97	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	99.40%
7) Chloroethane-d5	1.59	69	21651	4.99	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	99.80%
11) 1,1-Dichloroethene-d2	2.14	63	38373	3.81	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	76.20%
20) 2-Butanone-d5	3.97	46	64747	52.32	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	104.64%
24) Chloroform-d	4.41	84	52587	4.95	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.00%
26) 1,2-Dichloroethane-d4	5.09	65	27012	5.09	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.80%
32) Benzene-d6	5.10	84	101562	4.81	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.20%
36) 1,2-Dichloropropane-d6	6.13	67	32273	4.82	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	96.40%
41) Toluene-d8	7.36	98	90572	4.41	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.20%
43) trans-1,3-Dichloropropene-	7.67	79	10582	4.54	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	90.80%
46) 2-Hexanone-d5	8.14	63	41799	45.01	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	90.02%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	21955	4.68	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	93.60%
64) 1,2-Dichlorobenzene-d4	11.68	152	36349	5.09	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	101.80%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA V\Data\VV100818\
Data File : VV008162.D
Acq On : 08 Oct 2018 19:45
Operator : SY/MD
Sample : J5281-06
Misc : 25.0 mL/MSVOA V/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BEYL3

Quant Time: Oct 15 19:11:13 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR100518WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Oct 09 04:07:48 2018
Response via : Initial Calibration

