

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV100821\
 Data File : VV022694.D
 Acq On : 08 Oct 2021 11:45
 Operator : SY/MD
 Sample : M4027-04
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AT8

Quant Time: Oct 09 01:24:06 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR100721WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Oct 09 01:23:25 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	116788	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	117289	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.252	152	52350	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	35201	4.381	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	87.600%
7) Chloroethane-d5	1.568	69	32349	4.480	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	89.600%
11) 1,1-Dichloroethene-d2	2.104	63	51714	3.281	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	65.600%
20) 2-Butanone-d5	3.921	46	60825	28.870	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	57.740%
24) Chloroform-d	4.349	84	70751	4.273	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	85.400%
26) 1,2-Dichloroethane-d4	5.034	65	34153	4.250	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	85.000%
32) Benzene-d6	5.050	84	138212	3.958	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	79.200%
36) 1,2-Dichloropropane-d6	6.072	67	42256	4.087	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	81.800%
41) Toluene-d8	7.317	98	110221	3.652	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	73.000%
43) trans-1,3-Dichloroprop...	7.625	79	13407	3.909	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	78.200%
46) 2-Hexanone-d5	8.098	63	48787	35.293	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	70.580%
56) 1,1,2,2-Tetrachloroeth...	10.220	84	25112	3.585	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	71.600%
66) 1,2-Dichlorobenzene-d4	11.625	152	44905	4.626	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	92.600%
Target Compounds						
42) Toluene	7.400	91	3332	0.104	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV100821\
Data File : VV022694.D
Acq On : 08 Oct 2021 11:45
Operator : SY/MD
Sample : M4027-04
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AT8

Quant Time: Oct 09 01:24:06 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR100721WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Oct 09 01:23:25 2021
Response via : Initial Calibration

