

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV100920\
 Data File : VV018801.D
 Acq On : 09 Oct 2020 08:09
 Operator : SY/MD
 Sample : VV1009WBL02
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK74

Quant Time: Oct 09 09:24:13 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM100820WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Oct 09 07:31:27 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	325229	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	300870	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	144454	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	138887	47.00	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	94.00%
7) Chloroethane-d5	1.58	69	112491	47.29	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	94.58%
11) 1,1-Dichloroethene-d2	2.12	63	206947	36.32	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	72.64%
21) 2-Butanone-d5	3.91	46	187704	103.45	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	103.45%
24) Chloroform-d	4.38	84	244685	46.93	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.86%
26) 1,2-Dichloroethane-d4	5.06	65	178743	50.57	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.14%
32) Benzene-d6	5.07	84	502570	51.90	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	103.80%
36) 1,2-Dichloropropane-d6	6.09	67	168913	53.50	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	107.00%
41) Toluene-d8	7.34	98	416834	48.50	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.00%
43) trans-1,3-Dichloropropene-	7.64	79	70765	45.85	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	91.70%
47) 2-Hexanone-d5	8.11	63	96226	86.96	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	86.96%
57) 1,1,2,2-Tetrachloroethane-	10.23	84	152961	42.12	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	84.24%
64) 1,2-Dichlorobenzene-d4	11.65	152	169284	52.95	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	105.90%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV100920\
Data File : VV018801.D
Acq On : 09 Oct 2020 08:09
Operator : SY/MD
Sample : VV1009WBL02
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK74

Quant Time: Oct 09 09:24:13 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM100820WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Oct 09 07:31:27 2020
Response via : Initial Calibration

