

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV100920\
 Data File : VV018800.D
 Acq On : 09 Oct 2020 07:45
 Operator : SY/MD
 Sample : VV1009MBL01
 Misc : 5.0µ/5.0mL/100µL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK73

Quant Time: Oct 09 09:21:46 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM100820WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Oct 09 07:31:27 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	340513	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	308864	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	154272	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	143995	46.54	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	93.08%
7) Chloroethane-d5	1.58	69	118401	47.54	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	95.08%
11) 1,1-Dichloroethene-d2	2.12	63	216712	36.33	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	72.66%
21) 2-Butanone-d5	3.90	46	188046	98.98	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	98.98%
24) Chloroform-d	4.38	84	254967	46.71	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.42%
26) 1,2-Dichloroethane-d4	5.06	65	177719	48.03	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.06%
32) Benzene-d6	5.07	84	509993	51.31	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.62%
36) 1,2-Dichloropropane-d6	6.09	67	156572	48.31	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	96.62%
41) Toluene-d8	7.34	98	433851	49.17	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.34%
43) trans-1,3-Dichloropropene-	7.64	79	72177	45.56	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	91.12%
47) 2-Hexanone-d5	8.11	63	101271	89.15	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	89.15%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	164459	44.11	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	88.22%
64) 1,2-Dichlorobenzene-d4	11.65	152	176193	51.60	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	103.20%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV100920\
Data File : VV018800.D
Acq On : 09 Oct 2020 07:45
Operator : SY/MD
Sample : VV1009MBL01
Misc : 5.0µ/5.0mL/100µL/5.0mL/MSVOA_V/MEOH
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK73

Quant Time: Oct 09 09:21:46 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM100820WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Oct 09 07:31:27 2020
Response via : Initial Calibration

