

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW100923\
 Data File : VW032325.D
 Acq On : 09 Oct 2023 13:45
 Operator : SY/MD
 Sample : VIBLK322
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK322

Quant Time: Oct 10 04:47:21 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR092923WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Oct 07 05:00:49 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.539	114	105098	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.789	117	115203	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	52391	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	47389	5.544	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	110.800%
7) Chloroethane-d5	1.532	69	37376	5.387	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	107.800%
11) 1,1-Dichloroethene-d2	2.060	65	17665	4.900	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	98.000%
20) 2-Butanone-d5	3.815	46	94465	43.689	ug/L	0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	87.380%
24) Chloroform-d	4.256	84	71443	4.738	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.800%
26) 1,2-Dichloroethane-d4	4.950	65	34707	4.859	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.200%
32) Benzene-d6	4.966	84	142636	4.208	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.200%
36) 1,2-Dichloropropane-d6	5.995	67	45116	4.017	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	80.400%
41) Toluene-d8	7.249	98	134784	4.348	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.000%
43) trans-1,3-Dichloroprop...	7.561	79	16322	4.101	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	82.000%
46) 2-Hexanone-d5	8.030	63	81268	43.400	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	86.800%
56) 1,1,2,2-Tetrachloroeth...	10.156	84	34851	4.436	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	88.800%
66) 1,2-Dichlorobenzene-d4	11.564	152	48483	4.987	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	99.800%
Target Compounds						Qvalue
34) Trichloroethene	5.844	95	1425	0.129	ug/L	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV100923\
Data File : VV032325.D
Acq On : 09 Oct 2023 13:45
Operator : SY/MD
Sample : VIBLK322
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK322

Quant Time: Oct 10 04:47:21 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR092923WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Oct 07 05:00:49 2023
Response via : Initial Calibration

