

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV100923\
 Data File : VV032326.D
 Acq On : 09 Oct 2023 14:10
 Operator : SY/MD
 Sample : 04743-04
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Oct 10 01:19:20 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR092923WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Oct 07 05:00:49 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.538	114	99108	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.789	117	110761	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.191	152	50640	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	45615	5.659	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	113.200%
7) Chloroethane-d5	1.532	69	36328	5.552	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	111.000%
11) 1,1-Dichloroethene-d2	2.060	65	17525	5.155	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	103.200%
20) 2-Butanone-d5	3.815	46	85541	41.952	ug/L	0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	83.900%
24) Chloroform-d	4.256	84	68336	4.806	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.200%
26) 1,2-Dichloroethane-d4	4.950	65	33816	5.021	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.400%
32) Benzene-d6	4.966	84	136792	4.197	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.000%
36) 1,2-Dichloropropane-d6	5.995	67	42682	3.953	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	79.000%
41) Toluene-d8	7.249	98	130250	4.371	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.400%
43) trans-1,3-Dichloroprop...	7.561	79	15672	4.096	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	82.000%
46) 2-Hexanone-d5	8.034	63	74296	41.268	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	82.540%
56) 1,1,2,2-Tetrachloroeth...	10.159	84	32292	4.276	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	85.600%
66) 1,2-Dichlorobenzene-d4	11.567	152	45286	4.819	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	96.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV100923\
Data File : VV032326.D
Acq On : 09 Oct 2023 14:10
Operator : SY/MD
Sample : 04743-04
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Oct 10 01:19:20 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR092923WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Oct 07 05:00:49 2023
Response via : Initial Calibration

