

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW100923\
 Data File : VW032335.D
 Acq On : 09 Oct 2023 17:49
 Operator : SY/MD
 Sample : 04748-05DL 40X
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EX878DL

Quant Time: Oct 10 04:52:50 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR092923WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Oct 07 05:00:49 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.539	114	103997	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.789	117	112493	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	51487	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	45234	5.348	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	107.000%	
7) Chloroethane-d5	1.532	69	35725	5.203	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	104.000%	
11) 1,1-Dichloroethene-d2	2.060	65	17278	4.844	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	96.800%	
20) 2-Butanone-d5	3.828	46	90194	42.155	ug/L	0.03
Spiked Amount	50.000	Range 40 - 130	Recovery	=	84.300%	
24) Chloroform-d	4.256	84	70261	4.709	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	94.200%	
26) 1,2-Dichloroethane-d4	4.950	65	33611	4.756	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	95.200%	
32) Benzene-d6	4.963	84	137901	4.166	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	83.400%	
36) 1,2-Dichloropropane-d6	5.992	67	49345	4.499	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	90.000%	
41) Toluene-d8	7.249	98	129156	4.267	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	85.400%	
43) trans-1,3-Dichloroprop...	7.561	79	15290	3.934	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	78.600%	
46) 2-Hexanone-d5	8.034	63	83963	45.919	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	91.840%	
56) 1,1,2,2-Tetrachloroeth...	10.156	84	34918	4.552	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	91.000%	
66) 1,2-Dichlorobenzene-d4	11.564	152	47100	4.930	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	98.600%	
Target Compounds						
					Qvalue	
16) Methylene chloride	2.449	84	3987	0.401	ug/L	91
22) cis-1,2-Dichloroethene	3.825	96	5710	0.607	ug/L	98
34) Trichloroethene	5.834	95	76075	7.036	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV100923\
Data File : VV032335.D
Acq On : 09 Oct 2023 17:49
Operator : SY/MD
Sample : 04748-05DL 40X
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
EX878DL

Quant Time: Oct 10 04:52:50 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR092923WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Oct 07 05:00:49 2023
Response via : Initial Calibration

