

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW100923\
 Data File : VW032350.D
 Acq On : 10 Oct 2023 00:15
 Operator : SY/MD
 Sample : 04774-05
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 E05K6

Quant Time: Oct 10 06:01:12 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR092923WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Oct 10 05:42:01 2023
 Response via : Initial Calibration

Compound		R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards							
1) 1,4-Difluorobenzene		5.539	114	103757	5.000	ug/L	0.00
28) Chlorobenzene-d5		8.789	117	112853	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4		11.191	152	52143	5.000	ug/L	0.00
System Monitoring Compounds							
4) Vinyl Chloride-d3		1.278	65	42153	4.995	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	100.000%	
7) Chloroethane-d5		1.532	69	34442	5.028	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	100.600%	
11) 1,1-Dichloroethene-d2		2.060	65	16577	4.658	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	93.200%	
20) 2-Butanone-d5		3.812	46	89047	41.715	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	83.440%	
24) Chloroform-d		4.252	84	74071	4.976	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.600%	
26) 1,2-Dichloroethane-d4		4.950	65	32451	4.602	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.000%	
32) Benzene-d6		4.963	84	128905	3.882	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	77.600%	
36) 1,2-Dichloropropane-d6		5.992	67	46755	4.250	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	85.000%	
41) Toluene-d8		7.246	98	122313	4.028	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	80.600%	
43) trans-1,3-Dichloroprop...		7.561	79	15702	4.028	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	80.600%	
46) 2-Hexanone-d5		8.030	63	76388	41.643	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	83.280%	
56) 1,1,2,2-Tetrachloroeth...		10.159	84	33946	4.411	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	88.200%	
66) 1,2-Dichlorobenzene-d4		11.564	152	45172	4.668	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	93.400%	
Target Compounds							
						Qvalue	
13) Acetone		2.137	43	5471	3.325	ug/L	91
25) Chloroform		4.278	83	169901	10.561	ug/L	99
38) Bromodichloromethane		6.436	83	82650	6.546	ug/L	99
49) Dibromochloromethane		8.182	129	24058	3.083	ug/L	99
59) Bromoform		9.674	173	1415	0.373	ug/L #	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV100923\
Data File : VV032350.D
Acq On : 10 Oct 2023 00:15
Operator : SY/MD
Sample : 04774-05
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
E05K6

Quant Time: Oct 10 06:01:12 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR092923WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Oct 10 05:42:01 2023
Response via : Initial Calibration

