

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV100923\
 Data File : VV032319.D
 Acq On : 09 Oct 2023 11:21
 Operator : SY/MD
 Sample : 04708-15
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DCJS5

Quant Time: Oct 09 14:04:22 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR092923WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Oct 07 05:00:49 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.538	114	101736	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.789	117	108297	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.191	152	48612	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	45898	5.547	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	111.000%
7) Chloroethane-d5	1.532	69	35621	5.304	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	106.000%
11) 1,1-Dichloroethene-d2	2.059	65	17482	5.010	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	100.200%
20) 2-Butanone-d5	3.815	46	82140	39.244	ug/L	0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	78.480%
24) Chloroform-d	4.256	84	68348	4.683	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.600%
26) 1,2-Dichloroethane-d4	4.947	65	32519	4.703	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.000%
32) Benzene-d6	4.963	84	137985	4.330	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.600%
36) 1,2-Dichloropropane-d6	5.992	67	43024	4.075	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	81.600%
41) Toluene-d8	7.249	98	129984	4.461	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.200%
43) trans-1,3-Dichloroprop...	7.561	79	15795	4.222	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	84.400%
46) 2-Hexanone-d5	8.030	63	67495	38.343	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	76.680%
56) 1,1,2,2-Tetrachloroeth...	10.159	84	30707	4.158	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	83.200%
66) 1,2-Dichlorobenzene-d4	11.567	152	45549	5.049	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	101.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV100923\
Data File : VV032319.D
Acq On : 09 Oct 2023 11:21
Operator : SY/MD
Sample : 04708-15
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DCJS5

Quant Time: Oct 09 14:04:22 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR092923WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Oct 07 05:00:49 2023
Response via : Initial Calibration

