

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_V\DATA\VV101018\
 Data File : VV008240.D
 Acq On : 10 Oct 2018 18:26
 Operator : SY/MD
 Sample : J5315-06 50X
 Misc : 25.0 mL/MSVOA_V/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0J97

Quant Time: Oct 11 03:23:39 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR100518WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Oct 11 01:57:50 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	75745	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	74471	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	31275	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	17704	3.81	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	76.20%
7) Chloroethane-d5	1.58	69	15774	4.02	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	80.40%
11) 1,1-Dichloroethene-d2	2.13	63	25529	2.80	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	56.00%#
20) 2-Butanone-d5	3.98	46	66813	59.73	ug/L	0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	119.46%
24) Chloroform-d	4.41	84	44020	4.59	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.80%
26) 1,2-Dichloroethane-d4	5.09	65	22956	4.79	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.80%
32) Benzene-d6	5.10	84	79674	4.19	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	83.80%
36) 1,2-Dichloropropane-d6	6.12	67	27176	4.50	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	90.00%
41) Toluene-d8	7.36	98	65313	3.53	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	70.60%
43) trans-1,3-Dichloropropene-	7.67	79	8501	4.04	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	80.80%
46) 2-Hexanone-d5	8.14	63	43866	52.38	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	104.76%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	19154	4.52	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	90.40%
64) 1,2-Dichlorobenzene-d4	11.68	152	28458	4.79	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	95.80%

Target Compounds

						Qvalue
16) Methylene chloride	2.54	84	2498	0.50	ug/L	94
18) trans-1,2-Dichloroethene	2.79	96	1087	0.24	ug/L	94
22) cis-1,2-Dichloroethene	3.96	96	183189	36.26	ug/L	100
34) Trichloroethene	5.96	95	67460	12.30	ug/L	99
47) Tetrachloroethene	8.02	164	1174109	239.73	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_V\DATA\VV101018\
Data File : VV008240.D
Acq On : 10 Oct 2018 18:26
Operator : SY/MD
Sample : J5315-06 50X
Misc : 25.0 mL/MSVOA_V/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0J97

Quant Time: Oct 11 03:23:39 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR100518WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Oct 11 01:57:50 2018
Response via : Initial Calibration

