

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV101020\
 Data File : VV018837.D
 Acq On : 10 Oct 2020 17:00
 Operator : SY/MD
 Sample : L4287-20
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AK6

Quant Time: Oct 12 05:54:01 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM100820WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Oct 12 04:43:49 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	304925	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	284375	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	135808	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	90479	32.66	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	65.32%
7) Chloroethane-d5	1.58	69	84663	37.96	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	75.92%
11) 1,1-Dichloroethene-d2	2.12	63	140672	26.34	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	52.68%#
21) 2-Butanone-d5	3.91	46	175056	102.90	ug/L	0.02
Spiked Amount	100.000	Range	40 - 130	Recovery	=	102.90%
24) Chloroform-d	4.38	84	197696	40.44	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	80.88%
26) 1,2-Dichloroethane-d4	5.06	65	140362	42.36	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	84.72%
32) Benzene-d6	5.08	84	360070	39.34	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	78.68%
36) 1,2-Dichloropropane-d6	6.09	67	130930	43.88	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	87.76%
41) Toluene-d8	7.34	98	305502	37.60	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	75.20%#
43) trans-1,3-Dichloropropene-	7.64	79	51650	35.41	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	70.82%
47) 2-Hexanone-d5	8.11	63	84943	81.22	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	81.22%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	164820	48.01	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	96.02%
64) 1,2-Dichlorobenzene-d4	11.65	152	138750	46.16	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	92.32%

Target Compounds

					Ovalue
5) Vinyl chloride	1.32	62	20146	7.483	ug/L 96
13) Acetone	2.19	43	4114	2.901	ug/L 74
17) trans-1,2-Dichloroethene	2.78	96	3462	1.554	ug/L 85
20) cis-1,2-Dichloroethene	3.94	96	49954	21.219	ug/L 97
34) Trichloroethene	5.95	95	2264	0.872	ug/L 89
46) Tetrachloroethene	8.00	164	6350	3.337	ug/L 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV101020\
Data File : VV018837.D
Acq On : 10 Oct 2020 17:00
Operator : SY/MD
Sample : L4287-20
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AK6

Quant Time: Oct 12 05:54:01 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM100820WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Oct 12 04:43:49 2020
Response via : Initial Calibration

