

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_V\DATA\VV101118\
 Data File : VV008259.D
 Acq On : 11 Oct 2018 12:48
 Operator : SY/MD
 Sample : VSTD0.5
 Misc : 25.0 mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD0.5

Quant Time: Oct 11 15:33:27 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVS101118SIM.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Oct 11 16:30:59 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	30421	0.50	ug/L	0.00
5) Chlorobenzene-d5	8.89	117	27870	0.50	ug/L	0.00
10) 1,4-Dichlorobenzene-d4	11.29	152	11530	0.50	ug/L	0.00

System Monitoring Compounds						
2) Vinyl Chloride-d3	1.32	65	16327	0.50	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	10.00%#
4) 1,2-Dichloroethane-d4	5.09	65	11951	0.50	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	10.00%#
7) Toluene-d8	7.36	98	32974	0.50	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	10.00%#
9) 1,1,2,2-Tetrachloroethane-	10.25	84	9073	0.50	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	10.00%#
11) 1-2-dichlorobenzene-d4	11.67	152	12571	0.50	ug/L	0.00

Target Compounds						Qvalue
3) Vinyl chloride	1.32	62	16748	0.50	ug/L	98
6) Trichloroethene	5.96	95	13450	0.50	ug/L	88
8) 1,2-Dibromoethane	8.40	107	8391	0.50	ug/L #	97
12) 1,2,3-Trichloropropane	10.32	75	7318	0.50	ug/L #	100
13) 1,2-Dibromo-3-chloropropan	12.48	75	1150	0.50	ug/L #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_V\DATA\VV101118\
Data File : VV008259.D
Acq On : 11 Oct 2018 12:48
Operator : SY/MD
Sample : VSTD0.5
Misc : 25.0 mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.5

Quant Time: Oct 11 15:33:27 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVS101118SIM.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Oct 11 16:30:59 2018
Response via : Initial Calibration

