

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_V\DATA\VV101118\
 Data File : VV008260.D
 Acq On : 11 Oct 2018 13:24
 Operator : SY/MD
 Sample : VSTD001
 Misc : 25.0 mL/MSVOA_V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD001

Quant Time: Oct 11 15:34:36 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVS101118SIM.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Oct 11 16:30:59 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	29954	0.50	ug/L	0.00
5) Chlorobenzene-d5	8.89	117	27883	0.50	ug/L	0.00
10) 1,4-Dichlorobenzene-d4	11.29	152	11665	0.50	ug/L	0.00

System Monitoring Compounds						
2) Vinyl Chloride-d3	1.32	65	30410	0.95	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	19.00%#
4) 1,2-Dichloroethane-d4	5.09	65	22485	0.96	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	19.20%#
7) Toluene-d8	7.35	98	63755	0.97	ug/L	-0.02
Spiked Amount	5.000	Range	70 - 130	Recovery	=	19.40%#
9) 1,1,2,2-Tetrachloroethane-	10.25	84	16949	0.93	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	18.60%#
11) 1-2-dichlorobenzene-d4	11.67	152	24826	0.98	ug/L	0.00

Target Compounds						Qvalue
3) Vinyl chloride	1.32	62	30033	0.91	ug/L	98
6) Trichloroethene	5.96	95	24729	0.92	ug/L	86
8) 1,2-Dibromoethane	8.40	107	15853	0.94	ug/L	97
12) 1,2,3-Trichloropropane	10.32	75	13842	0.93	ug/L #	100
13) 1,2-Dibromo-3-chloropropan	12.48	75	2141	0.92	ug/L #	74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_V\DATA\VV101118\
Data File : VV008260.D
Acq On : 11 Oct 2018 13:24
Operator : SY/MD
Sample : VSTD001
Misc : 25.0 mL/MSVOA_V/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD001

Quant Time: Oct 11 15:34:36 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVS101118SIM.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Oct 11 16:30:59 2018
Response via : Initial Calibration

