

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_V\DATA\VV101118\
 Data File : VV008267.D
 Acq On : 11 Oct 2018 18:29
 Operator : SY/MD
 Sample : VSTDCCC0.5
 Misc : 25.0 mL/MSVOA_V/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD0.567

Quant Time: Oct 11 19:05:58 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVS101118SIM.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Oct 11 17:07:31 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	16336	0.50	ug/L	0.00
5) Chlorobenzene-d5	8.89	117	14546	0.50	ug/L	0.00
10) 1,4-Dichlorobenzene-d4	11.31	152	5447	0.50	ug/L	0.02

System Monitoring Compounds

2) Vinyl Chloride-d3	1.32	65	8924	0.50	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	100.00%
4) 1,2-Dichloroethane-d4	5.09	65	6696	0.55	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	110.00%
7) Toluene-d8	7.36	98	16904	0.52	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	104.00%
9) 1,1,2,2-Tetrachloroethane-	10.25	84	5211	0.59	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	118.00%
11) 1-2-dichlorobenzene-d4	11.67	152	5953	0.49	ug/L	0.00
Spiked Amount	0.500	Range	80 - 120	Recovery	=	98.00%

Target Compounds

						Qvalue
3) Vinyl chloride	1.32	62	9252	0.54	ug/L	100
6) Trichloroethene	5.96	95	7073	0.47	ug/L	88
8) 1,2-Dibromoethane	8.40	107	4808	0.58	ug/L #	97
12) 1,2,3-Trichloropropane	10.32	75	4259	0.64	ug/L #	100
13) 1,2-Dibromo-3-chloropropan	12.48	75	682	0.64	ug/L	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_V\DATA\VV101118\
Data File : VV008267.D
Acq On : 11 Oct 2018 18:29
Operator : SY/MD
Sample : VSTDCCC0.5
Misc : 25.0 mL/MSVOA_V/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.567

Quant Time: Oct 11 19:05:58 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVS101118SIM.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Oct 11 17:07:31 2018
Response via : Initial Calibration

