

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_V\DATA\VV101118\
 Data File : VV008261.D
 Acq On : 11 Oct 2018 13:53
 Operator : SY/MD
 Sample : VSTD002
 Misc : 25.0 mL/MSVOA_V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD002

Quant Time: Oct 11 15:34:58 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVS101118SIM.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Oct 11 16:30:59 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	29641	0.50	ug/L	0.00
5) Chlorobenzene-d5	8.89	117	27345	0.50	ug/L	0.00
10) 1,4-Dichlorobenzene-d4	11.30	152	11332	0.50	ug/L	0.00

System Monitoring Compounds

2) Vinyl Chloride-d3	1.32	65	62194	1.95	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	39.00%#
4) 1,2-Dichloroethane-d4	5.09	65	46013	1.98	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	39.60%#
7) Toluene-d8	7.35	98	133221	2.06	ug/L	-0.02
Spiked Amount	5.000	Range	70 - 130	Recovery	=	41.20%#
9) 1,1,2,2-Tetrachloroethane-	10.25	84	35309	1.98	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	39.60%#
11) 1-2-dichlorobenzene-d4	11.67	152	50917	2.06	ug/L	0.00

Target Compounds

						Qvalue
3) Vinyl chloride	1.32	62	60969	1.87	ug/L	97
6) Trichloroethene	5.96	95	49108	1.86	ug/L	88
8) 1,2-Dibromoethane	8.40	107	32105	1.95	ug/L	96
12) 1,2,3-Trichloropropane	10.32	75	28404	1.97	ug/L #	100
13) 1,2-Dibromo-3-chloropropan	12.48	75	4290	1.90	ug/L #	69

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_V\DATA\VV101118\
Data File : VV008261.D
Acq On : 11 Oct 2018 13:53
Operator : SY/MD
Sample : VSTD002
Misc : 25.0 mL/MSVOA_V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD002

Quant Time: Oct 11 15:34:58 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVS101118SIM.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Oct 11 16:30:59 2018
Response via : Initial Calibration

