

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW101121\
 Data File : VW022739.D
 Acq On : 11 Oct 2021 19:04
 Operator : SY/MD
 Sample : M4109-16
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 YATK0

Quant Time: Oct 12 03:43:29 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR100721WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Oct 12 03:20:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	114154	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	112974	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.252	152	51980	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	30975	3.944	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	78.800%	
7) Chloroethane-d5	1.568	69	32852	4.655	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	93.000%	
11) 1,1-Dichloroethene-d2	2.108	63	49463	3.211	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	64.200%	
20) 2-Butanone-d5	3.896	46	98134	47.653	ug/L	-0.02
Spiked Amount	50.000	Range 40 - 130	Recovery	=	95.300%	
24) Chloroform-d	4.349	84	71372	4.410	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	88.200%	
26) 1,2-Dichloroethane-d4	5.034	65	36594	4.658	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	93.200%	
32) Benzene-d6	5.050	84	139031	4.134	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	82.600%	
36) 1,2-Dichloropropane-d6	6.072	67	45789	4.597	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	92.000%	
41) Toluene-d8	7.317	98	115115	3.959	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	79.200%	
43) trans-1,3-Dichloroprop...	7.625	79	13728	4.155	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	83.200%	
46) 2-Hexanone-d5	8.092	63	59269	44.513	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	89.020%	
56) 1,1,2,2-Tetrachloroeth...	10.217	84	26833	3.977	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	79.600%	
66) 1,2-Dichlorobenzene-d4	11.625	152	48798	5.063	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	101.200%	
Target Compounds						
10) 1,1,2-Trichloro-1,2,2-...	2.114	101	988	0.134	ug/L #	52
13) Acetone	2.185	43	76226	66.380	ug/L #	43
25) Chloroform	4.378	83	8057	0.515	ug/L	98
30) Cyclohexane	4.677	56	863	0.072	ug/L #	82
34) Trichloroethene	5.918	95	21951	2.653	ug/L	96
47) Tetrachloroethene	7.976	164	41970	6.272	ug/L	97
51) Chlorobenzene	8.883	112	2177	0.107	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV101121\
Data File : VV022739.D
Acq On : 11 Oct 2021 19:04
Operator : SY/MD
Sample : M4109-16
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
YATK0

Quant Time: Oct 12 03:43:29 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR100721WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Oct 12 03:20:38 2021
Response via : Initial Calibration

