

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV101121\
 Data File : VV022741.D
 Acq On : 11 Oct 2021 19:52
 Operator : SY/MD
 Sample : M4109-08 500X
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 YATJ4

Quant Time: Oct 12 03:43:54 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR100721WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Oct 12 03:20:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	107568	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	109321	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	49081	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	11527	1.558	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	31.200%#	
7) Chloroethane-d5	1.564	69	32019	4.814	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	96.200%	
11) 1,1-Dichloroethene-d2	2.101	63	45173	3.112	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	62.200%	
20) 2-Butanone-d5	3.924	46	68027	35.056	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	70.120%	
24) Chloroform-d	4.346	84	74793	4.904	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	98.000%	
26) 1,2-Dichloroethane-d4	5.037	65	37801	5.107	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	102.200%	
32) Benzene-d6	5.047	84	143800	4.419	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	88.400%	
36) 1,2-Dichloropropane-d6	6.072	67	46166	4.790	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	95.800%	
41) Toluene-d8	7.317	98	114448	4.068	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	81.400%	
43) trans-1,3-Dichloroprop...	7.635	79	2348	0.734	ug/L	0.01
Spiked Amount	5.000	Range 55 - 130	Recovery	=	14.600%#	
46) 2-Hexanone-d5	8.098	63	61295	47.573	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	95.140%	
56) 1,1,2,2-Tetrachloroeth...	10.220	84	30101	4.610	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	92.200%	
66) 1,2-Dichlorobenzene-d4	11.625	152	47809	5.254	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	105.000%	
Target Compounds						
					Qvalue	
16) Methylene chloride	2.500	84	11255	1.258	ug/L	99
34) Trichloroethene	5.918	95	667	0.083	ug/L	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV101121\
Data File : VV022741.D
Acq On : 11 Oct 2021 19:52
Operator : SY/MD
Sample : M4109-08 500X
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
YATJ4

Quant Time: Oct 12 03:43:54 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR100721WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Oct 12 03:20:38 2021
Response via : Initial Calibration

