

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV101122\
 Data File : VV028513.D
 Acq On : 11 Oct 2022 11:05
 Operator : SY/MD
 Sample : VV1011WBL01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK278

Quant Time: Oct 12 00:27:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM100722WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Oct 12 00:26:27 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	5.612	114	285500	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.847	117	261584	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.242	152	132395	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	104495	47.434	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	94.860%
7) Chloroethane-d5	1.564	69	79897	45.740	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	91.480%
11) 1,1-Dichloroethene-d2	2.101	63	140594	37.499	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	75.000%
21) 2-Butanone-d5	3.889	46	145197	75.292	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	75.290%
24) Chloroform-d	4.342	84	210716	52.154	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	104.300%
26) 1,2-Dichloroethane-d4	5.024	65	147152	56.474	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	112.940%
32) Benzene-d6	5.043	84	444746	53.059	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	106.120%
36) 1,2-Dichloropropane-d6	6.063	67	134649	49.270	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	98.540%
41) Toluene-d8	7.310	98	403701	53.794	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	107.580%
43) trans-1,3-Dichloroprop...	7.616	79	71882	51.603	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	103.200%
47) 2-Hexanone-d5	8.085	63	131781	88.643	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	88.640%
56) 1,1,2,2-Tetrachloroeth...	10.207	84	174074	46.547	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	93.100%
66) 1,2-Dichlorobenzene-d4	11.615	152	147884	53.907	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	107.820%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV101122\
Data File : VV028513.D
Acq On : 11 Oct 2022 11:05
Operator : SY/MD
Sample : VV1011WBL01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK278

Quant Time: Oct 12 00:27:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM100722WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Oct 12 00:26:27 2022
Response via : Initial Calibration

