

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW101123\
 Data File : VW032383.D
 Acq On : 11 Oct 2023 11:07
 Operator : SY/MD
 Sample : 04708-11DL 10X
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DCJT3DL

Quant Time: Oct 12 03:03:15 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR092923WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Oct 12 02:54:58 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.539	114	106165	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.789	117	115286	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.191	152	53420	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	40060	4.640	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	92.800%	
7) Chloroethane-d5	1.532	69	32745	4.672	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	93.400%	
11) 1,1-Dichloroethene-d2	2.060	65	14575	4.003	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	80.000%	
20) 2-Butanone-d5	3.805	46	83804	38.369	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	76.740%	
24) Chloroform-d	4.256	84	65744	4.316	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	86.400%	
26) 1,2-Dichloroethane-d4	4.947	65	30591	4.240	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	84.800%	
32) Benzene-d6	4.966	84	124550	3.672	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	73.400%	
36) 1,2-Dichloropropane-d6	5.992	67	41004	3.648	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	73.000%	
41) Toluene-d8	7.249	98	116512	3.756	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	75.200%	
43) trans-1,3-Dichloroprop...	7.564	79	14518	3.645	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	73.000%	
46) 2-Hexanone-d5	8.030	63	74398	39.702	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	79.400%	
56) 1,1,2,2-Tetrachloroeth...	10.156	84	32758	4.167	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	83.400%	
66) 1,2-Dichlorobenzene-d4	11.564	152	43267	4.365	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	87.200%	
Target Compounds						
						Qvalue
16) Methylene chloride	2.452	84	3191	0.314	ug/L	88
25) Chloroform	4.285	83	11470	0.697	ug/L	99
34) Trichloroethene	5.834	95	67883	6.126	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV101123\
Data File : VV032383.D
Acq On : 11 Oct 2023 11:07
Operator : SY/MD
Sample : 04708-11DL 10X
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DCJT3DL

Quant Time: Oct 12 03:03:15 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR092923WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Oct 12 02:54:58 2023
Response via : Initial Calibration

